Physikalisch-Technische Bundesanstalt

Braunschweig und Berlin

Member State of OIML Germany



OIML Certificate N° R49-1/2006-DE1-08.02 Revision 1

OIML CERTIFICATE OF CONFORMITY

Issuing Authority

Name: Physikalisch-Technische Bundesanstalt Address: Bundesallee 100, 38116 Braunschweig

Person responsible: Dr. Gudrun Wendt

Applicant

Name: Elster Metering Limited

130 Camford Way

Address: Sundon Park

Luton, Bedfordshire LU3 3AN

United Kingdom

Manufacturer of the certified type is the applicant.

Identification of the certified type

Water meter intended for the metering of cold potable water Type: SM100, SM100E, SM100P or SM001, SM001E, SM001P

SM150, SM150E, SM150P SM250, SM250E, SM250P

Further characteristics see page 3

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R49-1 (2006): Metrological and technical requirements

R49-2 (2006): Test methods R49-3 (2006): Test report format

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

Physikalisch-Technische Bundesanstalt

OIML Certificate N° R49-1/2006-DE1-08.02 Revision 1

The conformity was established by the results of tests and examinations provided in the associated Report No. PTB-1.5-4036396, Revision 1 (93 pages) and Test Report No. PTB-1.5-4025664, Revision 1 (94 pages).

The Issuing Authority

The CIML Member

Dr. Gudrun Wendt Dr. Roman Schwartz

Head of Department Head of Division

Liquid Flow Mechanics and Acoustics

05.08.2008 05.08.2008

Important note: Apart from the mention of the Certificate's reference number and the name of the

OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report(s) is not permitted, although either

may be reproduced in full.

Physikalisch-Technische Bundesanstalt

OIML Certificate N° R49-1/2006-DE1-08.02 Revision 1

Identification of the certified pattern – page 1 continued

Metrology characteristics:

Model	9	SM150 (E,P)		SM250 (E,P)		
$Q_3 (m^3/h)$		2.5			4.0	
$Q_4 (m^3/h)$		3.125		5.0		
Q_2/Q_1		1.6		1.6		
$Q_1 (m^3/h)$	0.0156	0.0125	0.010	0.020	0.025	
$Q_2 (m^3/h)$	0.025	0.020	0.016	0.032	0.040	
Q_3/Q_1	160	200	250	200	160	
Lenght (mm)		110		190		
thread		G ¾" B		G 1" B		

Model	SM100 (SM100 (E,P) or SM001 (E,P)		
$Q_3 (\text{m}^3/\text{h})$		1.6		
$Q_4 (\text{m}^3/\text{h})$		2.0		
Q_2/Q_1		1.6		
$Q_1 (m^3/h)$	0.010	0.008	0.0064	
$Q_2(m^3/h)$	0.016	0.0128	0.01024	
Q_3/Q_1	160	200	250	
Lenght (mm)		110		
thread		G ¾" B		

Verification scale interval (m³)	0.00001		
Accuracy Class	2		
Temperature Class	T30		
Maximum admissible pressure (bar)	16		
Maximum admissible temperature (°C)	30		
Environmental Class	B and C		
Electromagnetic environment	Residential, Commercial and Light industrial use		

Installation details:

Connection type	Screw thread
Minimum straight length of outlet pipe	0 mm
Minimum straight length of inlet pipe	0 mm
Flow conditioner	none
Orientation limitations	none