

Physikalisch-Technische Bundesanstalt

Braunschweig und Berlin

Member State of OIML
Germany



OIML Certificate No.
R49/2006-DE1-08.02
Revision 5

OIML CERTIFICATE OF CONFORMITY

Issuing Authority

Name: Physikalisch-Technische Bundesanstalt
Address: Bundesallee 100, 38116 Braunschweig
Person responsible: Dr. Gudrun Wendt

Applicant

Name: Elster Metering Limited
Address: 130 Camford Way
Sundon Park
Luton, Bedfordshire
LU3 3AN
United Kingdom

Manufacturer of the certified type is the applicant and
Elster Metering LTD (Circle Ring Network)
No. 10 Jalan Jurutera U1/23, Section U1
Kawasan Perindustrial Hicom Glenmarie
40150 Shah Alam, Selangor Darul Ehsan
Malaysia

and

Elster s.r.o.
Nám. Dr. Alberta Schweitzera 194
91601 Stará Turá
Slovenská republika

Identification of the certified type

Water meter intended for the metering of cold potable water
Type: SM100, SM100E, SM100P or SM001, SM001E, SM001P
SM150, SM150E, SM150P
SM250, SM250E, SM250P
SM700, SM700E, SM700P

Further characteristics see page 3

Physikalisch-Technische Bundesanstalt

OIML Certificate No.
R49/2006-DE1-08.02
Revision 5

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R49-1 (2006): Metrological and technical requirements
R49-2 (2006): Test methods
R49-3 (2006): Test report format

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Report No.

PTB-1.5-4036396, Revision 3 (94 pages) and Test Report No.
PTB-1.5-4025664, Revision 5 (95 pages).

The Issuing Authority

Dr. G. Wendt
Head of Department

11.04.2013

The OIML Member

Dr. R. Schwartz
Head of Division

11.04.2013

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report(s) is not permitted, although either may be reproduced in full.

Physikalisch-Technische Bundesanstalt

OIML Certificate No.
R49/2006-DE1-08.02
Revision 5

Identification of the certified pattern – page 1 continued

Metrology characteristics:

Model	SM150 (E,P)			SM250 (E,P)		SM700 (E,P)	
Q ₃ (m ³ /h)	2.5			4.0		4.0	
Q ₄ (m ³ /h)	3.125			5.0		5.0	
Q ₂ /Q ₁	1.6			1.6		1.6	
Q ₁ (m ³ /h)	0.0156	0.0125	0.010	0.020	0.025	0.020	0.025
Q ₂ (m ³ /h)	0.025	0.020	0.016	0.032	0.040	0.032	0.040
Q ₃ /Q ₁	160	200	250	200	160	200	160
Lenght (mm)	110			190		190	
thread	G ¾" B			G 1" B		1" or 1¼" NPSM	

Model	SM100 (E,P) or SM001 (E,P)		
Q ₃ (m ³ /h)	1.6		
Q ₄ (m ³ /h)	2.0		
Q ₂ /Q ₁	1.6		
Q ₁ (m ³ /h)	0.010	0.008	0.0064
Q ₂ (m ³ /h)	0.016	0.0128	0.01024
Q ₃ /Q ₁	160	200	250
Lenght (mm)	110		
thread	G ¾" B		

Verification scale interval (m ³)	Arbitrary
Accuracy class	2
Temperature class	T30
Maximum admissible pressure (bar)	16
Maximum admissible temperature (°C)	30
Environmental class	B and C
Electromagnetic environment	Residential, Commercial and Light industrial use

Installation details:

Connection type	Screw thread
Minimum straight length of outlet pipe	0 mm
Minimum straight length of inlet pipe	0 mm
Flow conditioner	none
Orientation limitations	none