



## **OIML Member State Denmark**

## Annex I to OIML Certificate no. R49/2006-DK2-10.01

Task No. 117-29536.01 Page 1 of 1

Issuing authority

Name and address:

FORCE Certification A/S

Park Allé 345

DK - 2605 Brøndby

Denmark

Person responsible:

Lars Poder

Applicant/Manufacturer

Name and address:

Siemens A/S Flow Instruments

Nordborgvej 81 DK - 6430 Nordborg

Denmark

Identification of the

certified type:

Water meter, electromagnetic flow meter

Type designation:

MAG5100W DN50 - 300 with MAG8000CT

Corrections:

a) New address

The manufacturers name and address have been changed

from the name and address as follows:

Siemens AG

DE - 76181 Karlsruhe

Germany

b) New example of a label on page 5

The label has been changed because of the change of address as follows:

## **SIEMENS**

Order No.: Serial No.:

16 bar/232psi

SITRANS F M MAG 8000 CT E68203TC001AA2 MAWP (PS) at -20°C/-4°F: MAWP (PS) at 70°C/158°F: 7ME68203TC001AA2 123456H123 Lining: EPDM ASTM A 105 T.media min.:

16 bar/232psi 16 bar/232psi 0.1C/32°F 50°C/122°F EN 1092-1 PN16 T.media max.:

0200

Sensor material: Process connection: Meter orientation: All orientations Enclosure: IP68/NEMA 6P Fluid group: PED/L1

Software version: 3.09

Year of Manuf.: 2017 Q3: 160m3/h

Q3/Q1: 160 Q2/Q1: 1,6

Supply: Certificate no.:

Size DN: 100 (4 inch.)

Lithium battery inside R49/2006-DK2-10.1

Siemens AG, DE-76181 Karlsruhe

Made in France

**Issuing Authority:** 

FORCE Certification A/S, OIML Issuing Authority DK2

9 March 2018

Lars Poder

Certification Manager