



OIML Certificate of Conformity

OIML Member State
The Netherlands

Number R49-1/2003-NL1-06.01 revision 2
Project number 13200194
Page 1 of 4

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant	KROHNE Altometer Kerkeplaat 12, 3313 LC Dordrecht, The Netherlands
Manufacturer	KROHNE Altometer Kerkeplaat 12, 3313 LC Dordrecht, The Netherlands
Identification of the certified type	A water meter Type : OPTIFLUX x300C; OPTIFLUX x000F + IFC300y * Water meter intended for the metering of cold potable water, model "OPTIFLUX x300C; OPTIFLUX x000F + IFC300y*", class 1 and 2.
Characteristics	See page 2 and further

This Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

- R49-1/2006 (E):** Metrological and technical requirements
- R49-2/2006 (E):** Test methods
- R49-3/2006 (E):** Test Report format

Remarks *) With x being 1, 2, 4, 5 or 6 and with y being F or W.

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation identified above.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
7 May 2013

C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMI (see www.nmi.nl).





OIML Certificate of Conformity

OIML Member State
The Netherlands

Number R49-1/2003-NL1-06.01 revision 2
Project number 13200194
Page 2 of 4

The conformity was established by the results of tests and examinations provided in the associated OIML Test Reports:

- No. R49-1/2003-NL1-06.01 that includes 193 pages (including Annexes)
- No. NMI-12200395-01 that includes 9 pages (including Annexes)
- No. NMI-12200544-01 that includes 9 pages (including Annexes)
- No. NMI-12200544-02 that includes 9 pages (including Annexes)

Identification of the certified pattern

Water meter intended for metering cold potable water, based on an electromagnetic principle, designed to measure reverse flow, with straight inlet and outlet length, with no flow conditioner and equipped with an electronic calculating/indicating device.

Metrological characteristics:

Type	: OPTIFLUX x300C ^[1] , complete water meter OPTIFLUX x000F ^[1] + IFC300y ^[1] , combined water meter		
Maximum admissible pressure (bar)	: 16		
Min/max admissible temperature (°C)	: 0,1/50		
Orientation	: All positions		
Environmental class	: C		
Power supply	:		
Type	AC	DC	AC/DC
U _{max}	230 V	24 V	24 V
U _{min}	100 V	12 V	24 V
Frequency	50 – 60 Hz	-	AC: 50 - 60 Hz

^[1] With x being 1, 2, 4, 5 or 6 and with y being F or W.



OIML Certificate of Conformity

OIML Member State
The Netherlands

Number R49-1/2003-NL1-06.01 revision 2
Project number 13200194
Page 3 of 4

Meter size	DN25	DN40	DN50	DN65	DN80	DN100	DN125
Minimum flow rate Q1 (m ³ /h)	0,040	0,063	0,10	0,16	0,25	0,4	0,6
Transitional flow rate Q2 (m ³ /h)	0,064	0,10	0,16	0,25	0,40	0,6	1,0
Permanent flow rate Q3 (m ³ /h)	16	25	40	100	160	250	400
Overload flow rate Q4 (m ³ /h)	20	31,3	50	125	200	312,5	500
Nominal diameter (mm)	25	40	50	65	80	100	125
Accuracy Class	2			1			
Indicating range (m ³) ^{[2][4]}	99.999			999.999			
Verification scale interval (m ³) ^{[3][4]}	0,0001			0,001			

Meter size	DN150	DN200	DN250	DN300
Minimum flow rate Q1 (m ³ /h)	0,6	1,0	1,6	2,5
Transitional flow rate Q2 (m ³ /h)	1,0	1,6	2,6	4,0
Permanent flow rate Q3 (m ³ /h)	400	1000	1600	2500
Overload flow rate Q4 (m ³ /h)	500	1250	2000	3125
Nominal diameter (mm)	150	200	250	300
Accuracy Class	1			
Indicating range (m ³)	999.999	9.999.999		
Verification scale interval (m ³)	0,001			

^[2] The indicating range is programmable, stated here is the minimum indicating range.

^[3] The verification scale interval is programmable, stated here is the maximum value.

^[4] The display of the totalizer has 11 digits (including 1 digit for the decimal sign). The format of the totalizer must be such that demands of the indicating range and the verification scale interval are met.



OIML Certificate of Conformity

OIML Member State
The Netherlands

Number R49-1/2003-NL1-06.01 revision 2
Project number 13200194
Page 4 of 4

Meter size	DN500	DN800	DN900	DN1000	DN1200
Minimum flow rate Q1 (m ³ /h)	12,6	125	200	200	200
Transitional flow rate Q2 (m ³ /h)	20,2	200	320	320	320
Permanent flow rate Q3 (m ³ /h)	6300	10000	16000	16000	16000
Overload flow rate Q4 (m ³ /h)	7875	12500	20000	20000	20000
Nominal diameter (mm)	500	800	900	1200	1200
Accuracy Class	1				
Indicating range (m ³)	99.999.999				
Verification scale interval (m ³)	0,01				

Electronic revision number

The electronic revision number is used to lay down the software version and the hardware of the instrument. If either changes, the number is updated.

- Approved electronic revision numbers: 3.2.4; 3.2.6; 3.2.7; 3.3.0; 3.3.1; 3.3.2_ and 3.3.3_.

The electronic revision number is stored under menu items B3.6 and C5.1.6. See the manual in how to access the parameters.

Software specification:

The first approved version of the water meter didn't have an electronic revision number. The approved software version is identified as:

- Main software: 2.2.1. Menu items B3.3 and C5.1.5
- User interface: 3.1.0. Menu item B3.4 and C5.2.5

See the manual in how to access the parameters.

Revision History

This revision replaces the previous version.

Revision	Date	Change(s)
Initial	17 March 2006	-
1	28 June 2012	Addition of several extra sizes
2	7 May 2013	Addition of several meter sizes