



OIML CERTIFICATE OF CONFORMITY

Issuing Authority

Name Slovak Legal Metrology
Address Hviezdoslavova 31
974 01 Banská Bystrica, Slovakia
Person responsible Jaromír Markovič

Applicant

Name Ningbo Aimei Meter Manufacture Co., Ltd..
Address 68, West Town Road, Shangtian Town, Fenghua City
Zhejiang, 315511 P.R. of China

Manufacturer of the certified type

The applicant

Identification of the certified type

**Family of mechanical volumetric (rotary piston) water meters
for metering of cold water**

Type **PD-A..., PD-AP...**

For further characteristics see page 2

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):


R 49-1, edition 2006
Accuracy class 2

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.
This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report: N° 2013/MI-001/B013, that includes 99 pages.


The Issuing Authority
Ing. Jaromír Markovič, PhD.




The OIML Member
Dr. h.c. mult. prof. Ing. Jozef Mihok, PhD.

29 September 2013

29 September 2013

Important note:

Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.

1. Designation

Mechanical volumetric (rotary piston) water meters types *PD-A...* and *PD-AP...* intended for metering the volumes (consumption) of clean cold water in residential (households) and commercial use. It is installed into pipe lines in all installation positions.

2 Description

Essential parts of water meter:

- measuring chamber - included chamber, rotary piston and top plate with transmission shaft for connection of measuring part with register;
- mechanical register - digital drum with gearing mechanism for all figures, semi-dry-dial counter with glycerine; 8 digits indication;
- housing - *PD-A...* brass body, *PD-AP...* plastic body
- non return valve.

Non-essential parts of water meter:

- sieve in the inlet of the water meter

2.1 Metrological functions

- measuring the volume of water.

2.2 Software

- not applicable

2.3 Integrated equipment and functions

- pulse output with reed sensor switch (optional), K-factors 2 impulse/L.

3 Technical and metrological data

Type	-	<i>PD-A / PD-AP</i>							
Nominal diameter <i>DN</i>	mm	15				20			
Permanent flowrate Q_3	m ³ /h	2,5				4			
Minimum flowrate Q_1	m ³ /h	0,025	0,02	0,0156	0,0125	0,040	0,032	0,025	0,020
Transitional flowrate Q_2	m ³ /h	0,040	0,032	0,0250	0,020	0,064	0,0512	0,040	0,032
Overload flowrate Q_4	m ³ /h	3,125				5			
Ratio Q_3/Q_1	-	100	125	160	200	100	125	160	200
Ratio Q_2/Q_1	-	1,6							
Connection thread	-	G ¾ B				G 1B			
Construction length <i>L</i>	mm	115				165			
Installation orientation	-	all position							
Water temperature range Θ	°C	0,1 to 50 (T50)							
Maximum working pressure	bar	16							
Maximum permissible error in upper flow rates range $Q_2 \leq Q \leq Q_4$	%	± 2 (at $\Theta \leq 30^\circ\text{C}$) ± 3 (at $\Theta > 30^\circ\text{C}$)							
Maximum permissible error in lower	%	± 5							
Scale interval	m ³	0,000 02							
Capacity of calculator	m ³	9999							
Number of digits	-	0000,0000							

4 Interfaces and compatibility conditions

- not applicable

5 Marking and inscriptions

The following data shall be marked on the water meter:

- a) manufacturer's name or mark;
- b) type of water meter;
- c) year of production and serial number;
- d) flowrate Q_3 and ratio Q_3/Q_1 , (R);
- e) maximum working pressure;
- f) OIML Certificate of conformity number;
- g) Temperature class

The flow direction shall be marked on a water meter's body in form of an arrow.
Markings on water meter must comply with the requirements OIML R 49.

Manufacturer can use following trademarks on its water meters:

AIMEI

ASM



6 Security measures

The water meter shall be protected against unauthorized manipulation by one sealing mark securing the connection of housing of water meter against opening.

7 Documentation used for assessment purposes

- Test report No 2013/MI-001/B013;
- Manufacturer's technical documentation stored in folder *Ningbo_PD_00*.

8 Standards and regulations used for assessment purposes

- OIML R 49-1, edition 2006 (E);
- OIML R 49-2, edition 2006 (E);
- OIML R 49-3, edition 2006 (E).

