



Member State of OIML
United Kingdom of Great Britain
and Northern Ireland

OIML Certificate No R49/2013-GB1-17.01

OIML CERTIFICATE OF CONFORMITY

Issuing authority: NMO

Person responsible: Mannie Panesar – Head of Technical Services

Applicant: Arad Ltd

Dalia - Ramot Menashe

POB19239 Dalia Israel

Manufacturer: The applicant

Identification of the

certified pattern: A family of cold-water meters, designated Octave, utilising

an Ultrasonic measuring element and having a rated permanent flowrate Q₃ between 40 m³/h and 1000 m³/h.

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report) with the requirements of the following Recommendation of the International Organisation of Legal Metrology (OIML):

OIML R49 - Edition 2013(E) for accuracy class: 2

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

This certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the certificates reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.

This revision replaces previous versions of the certificate.

Issue Date: 3 April 2017

Marek Bokota Technical Manager

M. Bolista

For and on behalf of the Head of Technical Services



0135

The conformity was established by testing and examinations described in the associated Evaluation Report P02031 which includes 13 pages.

Characteristics of the instrument:

Meter Size (mm)	50	65	80	100	150	200
Q ₃ /Q ₁ (R)	500	500	500	500	500	500
Q_2/Q_1	1.6	1.6	1.6	1.6	1.6	1.6
Q ₁ Minimum flowrate (m ³ /hr)	0.08	0.08	0.125	0.20	0.5	0.8
Q ₂ Transitional flowrate (m ³ /hr)	0.128	0.128	0.2	0.32	0.8	1.28
Q ₃ Permanent flowrate (m ³ /hr)	40	40	63	100	250	400
Q ₄ Overload flowrate (m ³ /hr)	50	50	80	125	313	500

Table 1

Measuring principle: Ultrasonic

Accuracy Class: 2 Q_2/Q_1 1.6 Q_3/Q_1 500

Environmental class: T50 (0.1 °C to 50 °C) Environmental class: O (-25 °C to 55 °C)

Electromagnetic environment: E1 Maximum admissible temperature: 50 °C

Maximum admissible pressure: 1.6 Mpa (16 bar)

Pressure Loss Class 0.16 bar

Installation details

Connection type Flange
Minimum straight length of inlet pipe: U0
Minimum straight length of outlet pipe: D0
Flow conditioner (details if required): None

Mounting

Orientation: Can be installed in any position

Functionality

Checking Facilities: Measurement transducer, Calculator & Indicator

Checking Facilities Type:

Flow Measurement Direction: The meter may or may not measure reverse

flow depending on factory set-up - this should

be marked on the Data Label

Output options: Pulse Output Module

4-20mA Output Module AMR Output Module

Encoder Protocol Output Module

Software versions: Vers 3.15 or 3.16 or 3.17 or 4.00

Authorised Alternatives

50 mm Threaded Body Meter

Having a 50mm meter with the same technical specifications as described in table 1, but with a threaded connection replacing the flanges.

40 and 50 mm Plastic Body

As described in the certificate but having a plastic body meter with the following specifications.

Meter Size	Q₃/Q₁ (R)	Q ₃ m³/h	PRESSURE LOSS	U/D
40mm	250	40	Δp 16	U0,D0
50mm	500	40	Δp 16	U0,D0

Meter Size	40 mm	50 mm
Q ₃ /Q ₁ (R)	250	500
Q_2/Q_1	1.6	1.6
Q ₁ Minimum flowrate (m ³ /hr)	0.16	0.08
Q ₂ Transitional flowrate (m ³ /hr)	0.256	0.128
Q ₃ Permanent flowrate (m ³ /hr)	40	40
Q ₄ Overload flowrate (m ³ /hr)	50	50

250 mm Meter

Having the flanged design meter with 250 mm diameter having either a Q_3/Q_1 turndown ratio of 500 (R500) or 315 (R315), with the following related flowrates:

$Q_3/Q_1(R)$	315	500
Q_2/Q_1	1.6	1.6
Q_1 (m 3 /hr)	2	2
Q_2 (m 3 /hr)	3.2	3.2
Q_3 (m 3 /hr)	630	1000
Q_4 (m 3 /hr)	787.5	1250

300 mm Meter

As described for the 250 mm above but with flanges of 300 mm, all the related flowrates are identical.

Octave 2

As described in the certificate but having alternative internal electronics. The only visible difference to the meter is the software version number change as shown on the label.

CERTIFICATE HISTORY

ISSUE NO.	DATE	DESCRIPTION
R49/2013-GB1-17.01	3 April 2017	Certificate first issued.
-	-	No revisions have been issued.