

OIML Certificate of Conformity

OIML Member State

The Netherlands

Number R49/2013-NL1-15.01 Project number 13200116 Page 1 of 3

Issuing authority Person responsible:

NMi Certin B.V. C. Oosterman

Applicant and Manufacturer

Badger Meter Europa GmbH

Nürtinger Straße 76 72639 Neuffen

Germany

Identification of the

An ultrasonic water meter

certified type Type: E-Series

Characteristics

See page 2 and further

This Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML Type Evaluation Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R 49-1 (2013) "Water meters intended for the metering of cold potable water and hot water"

Accuracy class

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation identified above. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Type Evaluation Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

NMi Certin B.V., OIML Issuing Authority NL1

29 October 2015

C. Oosterman

Head Certification Board

NMi Certin B.V. Hugo de Grootplein 1 3314 EG Dordrecht the Netherlands T +31 78 6332332 certin@nmi.nl This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).







OIML Certificate of Conformity

OIML Member StateThe Netherlands

Number R49/2013-NL1-15.01 Project number 13200116 Page 2 of 3

The conformity was established by the results of tests and examinations provided in the associated report(s):

- No. NMi-13200116-01 dated 30 October 2015 that includes 39 pages;
- No. 130601023 dated 23 October 2015 that includes 74 pages.

Characteristics of the measuring instrument

In Table 1 the general characteristics of the measuring instrument are presented.

Table 2 gives an overview of the general characteristics of the family of instruments.

The construction of the measuring instrument is recorded in the Documentation folder no. T10555-1.

Table 1 General characteristics

Measuring principle	Ultrasonic flow metering		
Accuracy class	2		
Environmental class	M1 / O (installed outdoors)		
Electromagnetic environment + + + +	E2++++++++++++++++++		
Temperature range ambient	-25 °C / +55 °C		
Water temperature class	T50 (+0,1 °C / +50 °C)		
+Maximum admissible pressure (MAP) + +	1,6 MPa (16 bar) + + + + + + + + + + + +		
Reverse flow	The sensor is not intended to measure reverse flow		
Orientation	All positions (Horizontal, vertical or diagonal)		
Flow profile sensitivity class	U0 and D0 (0 x DN upstream and 0 x DN downstream)		
Pressure loss class + + + + + + + +	Δp 40 (0,40 bar) + + + + + + + + + + + + + + + + + + +		
Power supply	Non-replaceable battery (2,7 – 3,7 V)		
Software identification + + + + + +	Version number: 1.33 Checksum: 2995		

5



OIML Certificate of Conformity

OIML Member State The Netherlands

Number R49/2013-NL1-15.01 Project number 13200116 Page 3 of 3

Table 2 General characteristics of the family of instruments

Ø in- and		Flow rates [m³/h]				Ratio
Meter size	outlet [mm]	Minimum Q1	Transitional Q2	Permanent Q3	Overload Q4	Q3/Q1
DN20 (3/4")	+ + 20 + +	+ 0,01 + -	+ 0,016 +	+ + -4 + +	+ + 5 + +	+ 400 + 4
DN25 (1")	25	0,0625	0,1	10	12,5	160
DN32 (1 1/4")	32	0,0625	0,1	10	12,5	160
DN40 (1,5")	+ + 40 + +	+ 0,08 +	+ 0,128	+ + 16 + +	+ + 20 + +	+ +200 + -
+ DN50 (2") +	+ + 50 + +	0,1	0,16	+ + 25 + +	31,25	+ +250 + -

Meter size	Indicating range [m³]	Verification scale interval [m³]	
DN20; DN25; DN32	99999,9999	0,000001	
DN40; DN50-+++++	- + + + 999999,999 + + + +	+ + + + +0,00001+ + + + +	

01