



OIML Certificate

OIML Member StateThe Netherlands



Number R 49/2013-A-NL1-21.02 revision 1 Project number 3506421 Page 1 of 3

Issuing authority Person responsible:

NMi Certin B.V. M.Ph.D. Schmidt



Applicant

Badger Meter, Inc.

4545 West Brown Deer Road Milwaukee, WI 53224 United States of America

Manufacturer

Badger Meter Europe GmbH

Nürtinger Straße 76 72639 Neuffen Germany

Identification of the

certified type

An electromagnetic water meter

Type: M2000

Characteristics

See page 2 and further

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML Type Evaluation Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R 49-1 (2013) "Water meters intended for the metering of cold potable water and hot water"

Accuracy class 1

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation identified above. This Certificate does not bestow any form of legal international approval.

This certificate and supporting reports comply with the requirements of OIML-CS-PD-07 clause 6.2.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Type Evaluation Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

NMi Certin B.V., OIML Issuing Authority NL1 17 March 2022

Certification Board

NMi Certin B.V. Thijsseweg 11 2629 JA Delft The Netherlands T +31 88 636 2332 certin@nmi.nl www.nmi.nl This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.











OIML Certificate

OIML Member StateThe Netherlands



Number R 49/2013-A-NL1-21.02 revision 1 Project number 3506421 Page 2 of 3

The conformity was established by the results of tests and examinations provided in the associated reports:

- No. NMi-13200483-01 dated 24 November 2016 that includes 96 pages;
- No. 140901822 / M2000 dated 5 March 2016 that includes 58 pages.

Characteristics of the measuring instrument

In Table 1 the general characteristics of the measuring instrument are presented.

Tables 2 and 3 give an overview of the general characteristics of the family of instruments.

The construction of the measuring instrument is recorded in the Documentation folder no. T10970-1.

Table 1 General characteristics

Measuring principle			Electromagnetic flow metering						
Accuracy class				1					
Environi	Environmental class				M1 / O (installed outdoors)				
Electromagnetic environment				E2					
Temperature range ambient				-25 °C / +55 °C					
Water temperature class				T50 (+0,1 °C / +50 °C)					
Maximum admissible pressure (MAP)				1,6 MPa (16 bar)					
Orientation				All positions (Horizontal, vertical or diagonal)					
Flow profile sensitivity class				U0 and D0 (0 x DN upstream and 0 x DN downstream)					
Reverse flow				The sensor is intended to measure reverse flow					
Pressure	Pressure loss class				Δp 10 (0,10 bar)				
Power supply				AC version: 85 – 265 VAC (45 – 65 Hz) or DC version: 9 – 36 VDC (grounding mandatory)					
Software	identificati	on							
	Version		(Checksum for language English and:					
	version	German	Czech	Spanish	French	Russian	Swedish	Turkish	
	1.15	DBDE	A020	7909	B8D6	A351	F370	477E	
	1.18	6B41	6231	424E	83F6	74A1	DC9F	30AB	



 \mathbf{T}





OIML Member State

The Netherlands



Number R 49/2013-A-NL1-21.02 revision 1 Project number 3506421 Page 3 of 3

OIML Certificate

Table 2 General characteristics of the family of instruments



	Ø in- and		Ratio			
Meter size	outlet [mm]	Minimum Q1	Transitional Q2	Permanent Q3	Overload Q4	Q3/Q1
DN50	50	0,252	0,4032	63	78,75	250
DN65	65	0,4	0,64	100	125	250
DN80	80	0,64	1,024	160	200	250
DN100	100	1	1,6	250	312,5	250
DN125	125	1,6	2,56	400	500	250
DN150	150	2,52	4,032	630	787,5	250
DN200	200	4	6,4	1000	1250	250
DN250	250	6,4	10,24	1600	2000	250
DN300	300	10	16	2500	3125	250

Table 3 General characteristics of the indicating device

Meter size	Indicating range [m³]	Verification scale interval [m³]	
DN50	99 999	0,0001	
DN65; DN80; DN100; DN125; DN150	999 999	0,001	
DN200; DN250; DN300	9 999 999	0,01	

Production location

The water meter is produced at one of the following production locations:

Badger Meter Czech Republic s.r.o.
 Maříkova 2082/26, 621 00 Brno, Czech Republic

Certificate history:

This revision replaces the previous version.

Revision	Date	Description of the modification
Initial	11 January 2021	-
1	17 March 2022	Added production location.