

Member State
P.R.China

OIML Certificate No.
R51/2006-CN1-14.01

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: Office of OIML Affairs, General Administration of Quality Supervision, Inspection and Quarantine

Address: N0.9 Madiandonglu, Haidian District, Beijing, 100088 P.R.China

Applicant

Name: Zhongshan Multiweigh Packaging Machinery Co., Ltd.

Address: No.34, Zhenlian Road, Shangnan District, Fusha Town, Zhongshan, Guangdong, (528434) P.R.China

Manufacturer of the certified type (if the manufacturer is not the applicant)

Name:

Address:

Identification of the certified type

Automatic catchweighing instruments (R51)

Type	JW-C200	JW-C1000	JW-C2000	JW-C6000
Accuracy class	XIII(1)			
Max	200 g	1000 g	2000 g	6000 g
Min	25 g	50 g	50 g	200 g
Verification scale interval e	1 g	1 g	1 g	5 g
Number of verification scale interval n	200	1000	2000	1200
Max Speed	52 m/min			
Temperature range	0 °C~40 °C			
Humidity range	≤85 %RH			
Power supply	220 V, 50/60 Hz			
Software identification	P140705			

The numbers of the associated test reports and the included pages:

NIM-LSmc-1401 118 pages

This certificate attests the conformity of the above-mentioned type(represented by the samples identified in the associated test report) with the requirements of the following Recommendation(s) of the International Organization of Legal Metrology (OIML).

R: 51

Edition: 2006

For accuracy class (if applicable): XIII(1)

This certificate relates only to the metrological and technical characteristics of the type of the instrument concerned ,as covered by the relevant OIML International Recommendation(s).

This certificate does not bestow any form of legal international approval. The conformity was established by tests described in the associated test reports.

The issuing authority


Office of OIML Affairs

中国秘书处

September 25, 2014

2014

The CIML member



September 25,

