



Nederlands Meetinstituut

OIML Member state
The Netherlands

OIML Certificate N° R51/1996-NL1-06.02

Project number 510141

Page 1 of 2

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMi Certin B.V.
Address: Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
Person responsible: Ing. C. Oosterman

Applicant

Name: Ishida Co., Ltd.
Address: 959-1, Shimomagari
Ritto
Japan

Manufacturer of the certified type

Name: Ishida Co., Ltd.
Address: 959-1, Shimomagari
Ritto
Japan

Identification of certified type

Automatic catchweighing instrument
Family of type: DACS-Z-***-**

600 g \leq Max \leq 3000 g
Min \geq 5 g for class X(1)

e \geq 0.1 g

n \leq 6000 divisions

5 g \leq mass < 10 g maximum belt speed 50 m/min

10 g \leq mass < 200 g maximum belt speed 90 m/min

200 g \leq mass \leq Max maximum belt speed 120 m/min

Temperature range 0 °C / 40 °C

NMi Certin B.V.


Hugo de Grootplein 1, 3314 EG Dordrecht
P.O. Box 394, 3300 AJ Dordrecht, NL
phone +31 78 6332332
fax +31 78 6332309
certin@nmi.nl
www.nmi.nl

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi B.V. (see "Regulation objection and appeal against decisions of NMi B.V.")

NMi Certin B.V., chamber o.c. nr. 27.233.418

This document is issued under the provision that no responsibility is accepted and that the applicant gives warranty for each responsibility against third parties.

The notification of NMi Certin as Issuing Authority can be verified at www.oiml.org.



Nederlands Meetinstituut

OIML Member state
The Netherlands

OIML Certificate N° R51/1996-NL1-06.02

Project number 510141

Page 2 of 2

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the type-approval certificate and the description with number T6936 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R51


Edition 1996 (E)
for accuracy class X(1).

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Report:
N° R51/1996-NL1-06.02 that includes 48 pages.

The Issuing Authority
Ing. C. Oosterman
Manager Product Certification



19 April 2006

*
* *

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.