



OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMI Certin B.V.
Address: Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
Person responsible: Ing. C. Oosterman

Applicant

Name: Avery Weigh-Tronix
Address: Foundry Lane, Smetwick
West Midlands B66 2LP
United Kingdom

Manufacturer of the certified type

Name: Avery Weigh-Tronix
Address: Foundry Lane, Smetwick
West Midlands B66 2LP
United Kingdom

Identification of certified type

Automatic catchweighing instrument
Type: LS-3000, CW-3000 and GW-3000

Max \leq 60 kg
Min \geq 20 e for class Y(a)
Min \geq 50 e for class X(1)
e \geq 1 g
n \leq 3000 divisions
T \leq - (Max - e)
Maximum belt speed 38.1 m/min.
Weighing speed 80 packages per minute

Temperature range -10 °C / +40 °C



OIML Member state
The Netherlands

OIML Certificate N° R51/1996-NL1-06.04

Project number 607407

Page 2 of 2

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the type-approval certificate and the description with number T10040 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R51

Edition 1996 (E)

for accuracy class X(1) and Y(a).

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Report:

N° R51/1996-NL1-06.03A that includes 55 pages;

N° R51/1996-NL1-06.03B that includes 20 pages.

The Issuing Authority
Ing. C. Oosterman
Manager Product Certification

23 November 2006

*

**

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.