



OIML Member State  
P. R. China



OIML Certificate No.  
R51/2006-B-CN2-2018.01 Revision 1

## OIML CERTIFICATE

ISSUED UNDER SCHEME B

**OIML Issuing Authority** Name: National Institute of Metrology, China  
Address: No.18, Bei San Huan Dong Lu, Chaoyang Dist, Beijing, P.R.China  
Person responsible: Mr. Fang Xiang, Director

**Applicant** Name: Xuzhou WIKA Electronics Control Technology Co., Ltd.  
Address: No. 11 Baoliansi Road, Xuzhou Economic Development Zone, Jiangsu, P.R.C.

**Manufacturer** Name: Xuzhou WIKA Electronics Control Technology Co., Ltd.  
Address: No. 11 Baoliansi Road, Xuzhou Economic Development Zone, Jiangsu, P.R.C.

**Identification of the certified type** Automatic Catchweighing Instruments HEW61-057BC

This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R51-1 2006(E)**

For accuracy class: **Y (b)**

This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full.

OIML Issuing Authority: **National Institute of Metrology, China**

Date: 5 Jun 2019



No.18, Bei San Huan Dong Lu,  
Chaoyang Dist, Beijing,  
100029, P.R.China  
+86 10 64524602  
oimlia@nim.ac.cn  
www.nim.ac.cn

**Fang Xiang**  
Director  
National Institute of Metrology, China

The conformity was established by the results of tests and examinations provided in the associated OIML type evaluation report:

No. TER-R51/2006-CN2-2018.01 Revision 1 dated 5 Jun 2019 that includes 19 pages  
and the associated OIML test report:

No. LSmt2017-1255 dated 28 August 2018 that includes 48 pages

#### **OIML Certificate History**

<b>Revision No.</b>	<b>Date</b>	<b>Description of the modification</b>
0	8 October 2018	Initial issue
1	5 Jun 2019	Revise the name of the Applicant and Manufacturer to "Xuzhou WIKA Electronics Control Technology Co., Ltd." upon request of the applicant. Revise "HEW61-57BC" to "HEW61-057BC".

#### **Characteristics of the Instrument**

Type	HEW61-057BC
Accuracy class	Y (b)
Max	9000 kg
Min	1000 kg
Verification scale interval <i>e</i>	50 kg
Number of verification scale interval <i>n</i>	180
Rate of operation	10 loads/min
Temperature range	-10°C~ +40°C
Humidity range	≤85 %RH
Power supply	12 V/24 V
Software identification	907677_1V0