

Physikalisch-Technische Bundesanstalt

Braunschweig und Berlin

Member State of OIML
Germany



OIML Certificate N°
R51/2006-DE1-08.02

OIML CERTIFICATE OF CONFORMITY

Issuing Authority

Name: Physikalisch-Technische Bundesanstalt
Address: Bundesallee 100, 38116 Braunschweig
Person responsible: Dr. Panagiotis Zervos

Applicant

Name: TeleFrank GmbH
Address: Am Wildengrund 1
98553 Altendambach
Germany

Manufacturer of the certified type is the applicant.

Identification of the certified type Automatic catchweighing instrument (postal scale)
Type: LBHG

Further characteristics see page 2

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R51-1, edition 2006
for accuracy class Y(a)

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

Physikalisch-Technische Bundesanstalt

OIML Certificate N°
R51/2006-DE1-08.02

The conformity was established by the results of tests and examinations provided in the associated Report

No. 1.12-4029581 (12 pages)

and in the associated Test Report

No. 1.12-4029581/1 (56 pages)

The Issuing Authority

Dr. P. Zervos
Direktor und Professor

17.09.2008

The OIML Member

Dr. R. Schwartz
Direktor und Professor

17.09.2008

Identification of the pattern (continued)

Automatic electromechanical weighing instrument (postal scale) with a strain gauge load cell (SG-LC) and without lever work performed as single interval instrument.

Weighing mode	start-stop operation
Accuracy class	Y(a)
No n of verification scale intervals	≤ 3000
Minimum capacity Min	5e
Verification scale interval e	≥ 1 g
Maximum capacity Max	≤ 3 kg
Temperature range	+ 5°C / + 40°C

Tab. 1: Metrological characteristics of the weighing instrument LBHG

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report(s) is not permitted, although either may be reproduced in full.