

# Physikalisch-Technische Bundesanstalt

Braunschweig und Berlin

Member State of OIML  
Germany



OIML Certificate N°  
**R51/2006-DE1-08.03**

## OIML CERTIFICATE OF CONFORMITY

### Issuing Authority

Name: Physikalisch-Technische Bundesanstalt  
Address: Bundesallee 100, 38116 Braunschweig  
Person responsible: Dr. Panagiotis Zervos

### Applicant

Name: LEICH und MEHL und Co. GmbH  
Address: Porschestraße 7  
71394 Kernen im Remstal  
GERMANY

Manufacturer of the certified type is the applicant.

**Identification of the certified type** Automatic catchweighing instrument  
Type: PAW2000-G

Further characteristics see page 2

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**R51-1**, edition 2006  
for accuracy classes XIII(1) and Y(a)

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

# Physikalisch-Technische Bundesanstalt

OIML Certificate N°  
**R51/2006-DE1-08.03**

The conformity was established by the results of tests and examinations provided in the associated Report

No. 1.12-4036537 (16 pages)

and in the associated Test Reports

No. 1.12-4036537/1 (70 pages) and

No. 1.12-4036537/2 (13 pages)

## The Issuing Authority

Dr. P. Zervos  
Direktor und Professor

05.08.2008

## The CIML Member

Dr. R. Schwartz  
Direktor und Professor

05.08.2008

## Identification of the pattern (continued)

Automatic electromechanical weighing instrument as

- catchweigher,
  - weigh price labeller,
  - weigh labeller or
  - checkweigher,
- equipped
- with electro-dynamic force compensation load cell (EFC-LC) with lever work and performed as
  - single or multi interval instrument.

# Physikalisch-Technische Bundesanstalt

OIML Certificate N°  
R51/2006-DE1-08.03

Weighing mode	Static weighing		Dynamic weighing	
	XIII(1)	Y(a)	XIII(1)	Y(a)
Verification scale interval e	e ≥ 1 g		e ≥ 2 g	
Ratio between verification scale intervals	$\frac{e_{i+1}}{e_i} < 3$			
Number n of verification scale intervals	≤ 6000 ≤ 2 • 6000		≤ 6000 ≤ 6000 / 4000	
Maximum load Max	≤ 60 kg			
Minimum load Min	≥ 20e <sup>1)</sup>		≥ 50 g <sup>2)</sup>	≥ 40 g (= 20e) <sup>2)</sup>
Temperature range	0 °C / + 40 °C			
Maximum belt speed	≤ 0,9 m/s			
Put through	≤ 80 pcs/min			
Tare	≤ 0,25 • Max <sup>3)</sup>			

Tab 1: Technical data of the weighing instrument PAW 2000 (EFC-LC of the type LB60)

- 1) This value is valid for static weighing. The minimum capacity in case of static weighing for category Y depends on the specification according to R51-1:2006 No 2.2.2. In that case it may be applied to corresponding category X in the same way in order to achieve smaller minimum capacities.
- 2) This value is valid for dynamic weighing. For weighing instruments of the category X the minimum loads depend on the used digital LC and for weighing instruments of the category Y additionally on the specification under 2). In case of category Y the greater value has to be taken. Greater minimum loads may also result from the metrological test of the LMC.
- 3) This specification applies only to weighing instruments with parameters being generally valid over the complete weighing range for all products and weighing in motion (dynamic weighing). For single products having especially their own product data storage and/or their own product-specific parameters greater tare loads are allowed if they are approved by the LMC. A tare load ≤ - Max is allowed for start stop operation (static weighing).

**Important note:** Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report(s) is not permitted, although either may be reproduced in full.