



Physikalisch-Technische Bundesanstalt
Braunschweig und Berlin

OIML Member State
Germany

OIML Certificate No.
R51/2006-A-DE1-2022.02

OIML CERTIFICATE ISSUED UNDER SCHEME A

OIML Issuing Authority

Name: Physikalisch-Technische Bundesanstalt,
Conformity Assessment Body
Address: Bundesallee 100, 38116 Braunschweig, GERMANY
Person responsible: Dr.-Ing. Prof. h. c. Frank Härtig

Applicant

Name: BROSA GmbH
Address: Dr.-Klein-Str. 1, 88069 Tett nang, Germany

Manufacturer

Name: BROSA GmbH
Address: Dr.-Klein-Str. 1, 88069 Tett nang, Germany

Identification of the certified type *(the detailed characteristics will be defined in the additional pages)*

Automatic catchweigher
Type: 0656 FlexLim (Safe)

Designation of the module *(if applicable)*

Not applicable

This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 51

Edition (year): 2006

For accuracy class (if applicable): Y(b)

This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated OIML type evaluation report:

No. 1.12-4105847 dated 09.06.2022 that includes 11 pages

The technical documentation relating to the identified type is contained in documentation file:

No. ZDS-R51/2006-A-DE1-2022.02 dated 09.06.2022 that includes 3 pages

OIML Certificate History

| Revision No. | Date | Description of the modification |
|--------------|------------|---------------------------------|
| --- | 09.06.2022 | Initial issuing |
| | | |
| | | |
| | | |

Identification, signature and stamp

The Issuing Authority


Daniela Marencke

Member of Conformity Assessment Body

Date: 09.06.2022



Table 1: Essential data

| Operating mode | | Operation in motion |
|------------------------------------|-------|---------------------|
| Accuracy class | | Y(b) |
| Minimum load Min | t | ≥ 1 |
| Verification scale interval e | t | ≥ 0.2 |
| Maximum capacity Max | t | ≤ 40 |
| Number n of scale intervals | | ≤ 200 |
| Tare-balancing range (subtractive) | • Max | $\leq 25 \%$ |
| Zero setting interval | min | ≤ 1000 |
| Warm-up time | min | ≤ 0 |

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full.