



## OIML CERTIFICATE OF CONFORMITY

### Issuing authority

Name: NMI Certin B.V.  
Address: Hugo de Grootplein 1  
3314 EG Dordrecht  
The Netherlands  
Person responsible: Ing. C. Oosterman

### Applicant

Name: Teraoka Seiko Co. Ltd  
Address: 13-12 Kugahara, 5-Chome  
Ohta-Ku, Tokyo 146-8580  
Japan

### Manufacturer of the certified type

Name: Teraoka Seiko Co. Ltd  
Address: 13-12 Kugahara, 5-Chome  
Ohta-Ku, Tokyo 146-8580  
Japan

### Identification of certified type

Automatic catchweighing instrument  
Type: AW-4600CPR-...

$6 \text{ kg} \leq \text{Max} \leq 15 \text{ kg}$   
 $\text{Min} \geq 20 \text{ e}$  for class Y(a)  
 $e \geq 1 \text{ g}$   
 $n \leq 3000$  divisions (per partial weighing range for multi-interval instruments)  
Maximum of two partial weighing ranges  
 $T \leq -\text{Max}_1$   
Weighing speed 30 packages per minute.

Temperature range  $-10 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$



**OIML Member state  
The Netherlands**

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the type-examination certificate and the description with number T10172 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**R51**

Edition 2006 (E)  
for accuracy class Y(a).

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Report:

N° R76/1992-NL1-04.40 that includes 44 pages;

N° R51/2006-NL1-08.01 that includes 18 pages;

N° 08023867 that includes 14 pages;

The Issuing Authority  
Ing. C. Oosterman  
Manager Product Certification

28 November 2008

\*  
\* \*

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.