



OIML Member state  
The Netherlands

## OIML CERTIFICATE OF CONFORMITY

### Issuing authority

Name: NMI Certin B.V.  
Address: Hugo de Grootplein 1  
3314 EG Dordrecht  
The Netherlands

### Applicant

Name: Teraoka Seiko Co. Ltd  
Address: 13-12 Kugahara, 5-Chome  
Ohta-Ku, Tokyo 146-8580  
Japan

### Manufacturer of the certified type

Name: Teraoka Seiko Co. Ltd  
Address: 13-12 Kugahara, 5-Chome  
Ohta-Ku, Tokyo 146-8580  
Japan

### Identification of certified type

Automatic catchweighing instrument  
Type: AW-4600CPR-...

$6 \text{ kg} \leq \text{Max} \leq 15 \text{ kg}$   
 $\text{Min} \geq 20 \text{ e}$  for class Y(a)  
 $e \geq 1 \text{ g}$   
 $n \leq 3000$  divisions (per partial weighing range for multi-interval instruments)  
Maximum of two partial weighing ranges  
 $T \leq -\text{Max}_1$   
Weighing speed 30 packages per minute.

Temperature range  $-10 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$

**NMI Certin BV**  
Hugo de Grootplein 1, 3314 EG Dordrecht  
PO Box 394, 3300 AJ Dordrecht, NL  
T +31 78 6332332  
F +31 78 6332309  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision  
that no liability is accepted and that the  
applicant shall indemnify third-party liability.

Reproduction of the complete document is  
permitted.  
The notification of NMI Certin as Issuing  
Authority can be verified at [www.oiml.org](http://www.oiml.org).





**OIML Member state**  
**The Netherlands**

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the type-examination certificate and the description with number T10172 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**R51**  
Edition 2006 (E)  
for accuracy class Y(a).

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Report:

- N° R76/1992-NL1-04.40 that includes 44 pages;
- N° R51/2006-NL1-08.01 that includes 18 pages;
- N° 08023867 that includes 14 pages;
- N° R76/1992-NL1-10.10 that includes 34 pages;
- N° R76/1992-NL1-10.11 that includes 26 pages.

The Issuing Authority NL1  
NMI Certin, 22 March 2010

  
C. Oosterman  
Head of Certification Board

\*  
\* \*

**Important note:** Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.