



OIML Certificate of Conformity

OIML Member State
The Netherlands

Number R51/2006-NL1-10.02
Project number 9200406
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant	Martini s.r.l. Processing & Packaging Machines Via Borgo 21 IT 35015 Gallieria Venta Padova Italy
Manufacturer	Martini s.r.l. Processing & Packaging Machines Via Borgo 21 IT 35015 Gallieria Venta Padova Italy
Identification of the certified type	An Automatic catchweighing instrument Type : MCW
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R51 - Edition 2006 (E) for accuracy class XIII(1)

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
28 May 2010



C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).





OIML Certificate of Conformity

OIML Member State
The Netherlands

Number R51/2006-NL1-10.02
Project number 9200406
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

Number R51/2006-NL1-10.02 dated 28 May 2010 that includes 43 pages.

Characteristics of the automatic catchweighing instrument

Max	:	Max \geq 3 kg
Min	:	\geq 50 g for class XIII(1)
Verification scale interval	:	$e \geq$ 1 g
Maximum number of scale intervals	:	$n \leq$ 3000 divisions
Preset tare	:	T \leq -Max
Maximum belt speed	:	75 m/min.
Weighing speed	:	150 packages per minute
Temperature range	:	+5 °C / +40 °C, non-condensing