



OIML Certificate of Conformity

OIML Member State
The Netherlands

Number R51/2006-NL1-11.03
Project number 11200666
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant	Teraoka Seiko Co., Ltd. 13-12 Kugahara 5-Chome Ohta-Ku Tokyo 146-8580 Japan
Manufacturer	Teraoka Seiko Co., Ltd. 13-12 Kugahara 5-Chome Ohta-Ku Tokyo 146-8580 Japan
Identification of the certified type	An Automatic catchweighing instrument Type : LI-4600
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R51 - Edition 2006 (E) for accuracy class Y(a)

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
15 November 2011


C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).





OIML Certificate of Conformity

OIML Member State
The Netherlands

Number R51/2006-NL1-11.03
Project number 11200666
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- Number R76/1992-NL1-04.40 dated 28 December 2004 that includes 44 pages;
- Number R51/2006-NL1-08.01 dated 21 November 2008 that includes 18 pages;
- Number 08023867 dated 18 February 2008 that includes 14 pages;
- Number NMI-11200666 dated 15 November 2011 that includes 22 Pages;
- Number 11071135 dated 26 August 2011 that includes 26 Pages.

Characteristics of the automatic catchweighing instrument

- Max : $\text{Max} \leq 4 \text{ kg}$
- Min : $\geq 20 \text{ e}$ for class Y(a)
- Verification scale interval : $e \geq 1 \text{ g}$
- Maximum number of scale intervals : $n \leq 2000$ divisions
- Tare : $T \leq -\text{Max}$
- Maximum belt speed : 32,6 m/min.
- Temperature range : $- 10 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$