

OIML Certificate of Conformity

OIML Member State The Netherlands Number R51/2006-NL1-16.01 Project number 15200633 Page 1 of 3

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman									
Applicant and Manufacturer	Premier Tech 1, Avenue Premier RIVIERE-DU-LOUP (QUEBEC) Canada G5R 6C1									
Identification of the certified type	An Automatic catchweighing instrument Type : DYN.CHECK MP									
+ + + + + + + + + +	Type : DYN.CHECK MP									
Characteristics	See next page									
+ identified in the OIML	the conformity of the above identified Type (represented by the sample(s) Test Report) with the requirements of the following Recommendation of the tion of Legal Metrology (OIML):									
	OIML R 51 - Edition 2006 (E) for accuracy class XIII(1) or Y(a)									
This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.										
<i>Important note:</i> Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.										
+ Issuing Authority	NMi Certin B.V., OIML Issuing Authority NL1									
+ + + + + + + + +	19 April 2016 + + + + + + + + + + + + + + + + + + +									
* * * * * * * * *	C. Øosterman + + + + + + + + + + + + + + + + + + +									
* * * * * * * * * *	Head Certification Board + + + + + + + + + + + + + + + + + + +									
+ NMi Certin B.V. + + + + + + + + + + + + + + + + + +	This document is issued under the Parties concerned can provision that no liability is lodge objection against									
3314 EG Dordrecht	accepted and that the applicant this decision, within six $(1 - 1)$ $(2 - 1)$									
+ the Netherlands T +31 78 6332332	shall indemnify third-party liability. weeks after the date of submission, to the									
certin@nmi.nl www.nmi.nl	The notification of NMi Certin B.V. general manager of NMi as Issuing Authority can be verified (see www.nmi.nl). at www.oiml.org									



OIML Certificate of Conformity

OIML Member State The Netherlands Number R51/2006-NL1-16.01 Project number 15200633 Page 2 of 3

The conformity was established by the results of tests and examinations provided in the associated	
OIML Test Report(s): + + + + + + + + + + + + + + + + + + +	
- No. NMi-13200671-01 dated 24 July 2014 that includes 57 pages;	
- No. NMi-13200671-02 dated 24 July 2014 that includes 10 pages;	
 No. NMi-13200671-03 dated 24 July 2014 that includes 41 pages; 	
 No. NMi-13200671-04 dated 24 July 2014 that includes 7 pages; 	
- No. 1238 dated 12 December 2012 that includes 22 pages;	
- No. 1.12-4039210-2 dated 30 June 2009 that includes 21 pages;	
- No. NMi-13200670-01 dated 24 July 2014 that includes 23 pages. + + + + + + + + + + + + + +	

Characteristics of the automatic catchweighing instrument

Destined to be used as	catchweigher, checkweigher, weigh labeller, weigh price labeller					
Weighing range(s)	Single interval Multi-interval Multiple range					
Maximum load transport system speed	75 m/min.					
Electromagnetic environment class	E2					
temperature range	- + + + + + + -10 °C	/ +40 °C + + + + + + + +				
Climatic humidity	• • • • • Non-condensing					
intended location	Closed					
Maximum capacity	Max ≥ 30 kg					
Verification scale interval	• • • • • • • • • • • • • • • • • • •	10 g + + + + + + + +				
Accuracy class	Y(a)	XIII(1) actual accuracy class determined at the time of putting into use				
Catchweig	her weighing dynamically					
Minimum	+ + +Min ≥ 100 e + + ·	+ + Min ≥ 300 e + +				
capacity multi-interval multiple range	$Min \ge 100 e_1$	$Min \ge 300 e_1$				
Maximum number of scale intervals	ial weighing range)					
Maximum number of partial weighing ranges	* * * * * * * * *	2 + + + + + + + + +				
Catchweigher w	eighing statically or Min ≥ 2	0 kg				
Minimum capacity	Min	≥ 20 e				
Maximum number of scale intervals	$n \leq$ the number of scale intervals mentioned in the certificates involved					
The instrument is weighing dynamically o	or statically during automat	ic operation.				



OIML Certificate of Conformity

OIML Member State The Netherlands Number R51/2006-NL1-16.01 Project number 15200633 Page 3 of 3

* Producers: * * * * * * * * * * * * * * * * * * *				
+ + + + + + + + + + + + + + + + + + + +				
Premier Tech Chronos				
Via G. Di Vittorio, 78				
Lemignano di Collecchio				
Parma				
+ 43044 + + + + + + + + + + + + + + + + + +				
+ ITALY + + + + + + + + + + + + + + + + + + +				
Chronos Richardson India Private Limited				
Plot No. 5E, Sector 4				
Ballabgarh				
Haryana				
+ 121004 + + + + + + + + + + + + + + + + + +				
Premier Tech Chronos (Thailand)				
93 Moo, 9 Bangna-Trad Km. 19 Road Tambol Bangchalong, Amphor Bangplee				
Samutprakarn 10540				
* Bangkok * * * * * * * * * * * * * * * * * *				
+ 10540 + + + + + + + + + + + + + + + + + + +				
+ THAILAND+ + + + + + + + + + + + + + + + + + +				
Premier Tech Chronos (Bag Treatment Holland B.V.)				
Meerheide 40				
Eersel				
Brabant-Septentrional				
+ 5521 DZ + + + + + + + + + + + + + + + + + +				
NETHERLANDS				
Premier Tech Chronos GmbH				
Löhestr. 18				
* Hennef* * * * * * * * * * * * * * * * * * *				
+ D-53773 + + + + + + + + + + + + + + + + + +				
+ GERMANY + + + + + + + + + + + + + + + + + + +				
Premier Tech Chronos				
Unit 1, Centurion Business Centre				
Blenheim Industrial Estate				
+ Nottingham + + + + + + + + + + + + + + + + + + +				
+ Basford + + + + + + + + + + + + + + + + + + +				
+ NG6 8WN + + + + + + + + + + + + + + + + + + +				
ENGLAND				