

**OIML Member State**  
The Netherlands

Number R51/2006-B-NL1-18.02  
Project number 1901630  
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant and Manufacturer	Bizerba SE & Co. KG Wilhelm-Kraut-Straße 65 72336 Balingen Germany
Identification of the certified type	An <b>Automatic catchweighing instrument</b> Type : WP ...
Characteristics	See next page

This OIML Certificate is issued under scheme B.


This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R 51** - Edition 2006 (E) for accuracy class [XIII(1), XIII(2), Y(a) of Y(b)]

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.  
This Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**  
13 July 2018



C. Oosterman  
Head Certification Board

NMi Certin B.V.  
Hugo de Grootplein 1  
3314 EG Dordrecht  
the Netherlands  
T +31 78 6332332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)



**OIML Member State**  
The Netherlands

Number R51/2006-B-NL1-18.02  
Project number 1901630  
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Type Evaluation Reports:

- No. NMI-1901630-01 dated 13 July 2018 includes 54 pages;
- No. NMI-1901630-02 dated 13 July 2018 includes 16 pages.

### Characteristics of the automatic catchweighing instrument

Destined to be used as	catchweigher, weight labeller, weight/price labeller			
Accuracy class	Y(a), Y(b)			
Maximum capacity	3 kg ≤ Max ≤ 15 kg			
Minimum capacity	Min ≥ 20 e for class Y(a) Min ≥ 10 e for class Y(b)			
Verification scale interval	e ≥ 1 g			
Weighing ranges	Single interval Multi-interval			
Maximum number of scale intervals (single interval)	n ≤ 3000 divisions			
Maximum number of scale intervals (multi-interval)	n ≤ 3000 divisions (per partial weighing range)			
Maximum number of partial weighing ranges	2			
Tare	T ≤ -0,2 Max			
Electromagnetic environment class	E2			
Climatic environment	temperature range	-10 °C / +40 °C		
	humidity	non-condensing		
	intended location	closed		
Power supply voltage	400/230 V AC 50 Hz			
Software identification for program modules	RetailPowerScale Version: 012		Scale OEM Module Version: 011	
	Windows	Linux	Windows	Linux
	6914	7130	3362	1418
Software identification for A/D board	047195xxxx (x= 0...9)			