

**OIML Member State**  
The Netherlands

Number R51/2006-B-NL1-18.03  
Project number 1901553  
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant and Manufacturer	VWS Nederland B.V. Hoogschaijksestraat 25 5374 EC Schaijk The Netherlands
Identification of the certified type	An <b>Automatic catchweighing instrument</b> Type : Octopus Max
Characteristics	See next page

This OIML Certificate is issued under scheme B.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R 51** - Edition 2006 (E) for accuracy class XIII(x), Y(a)

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.

This Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**  
9 April 2018



C. Oosterman  
Head Certification Board

NMi Certin B.V.  
Hugo de Grootplein 1  
3314 EG Dordrecht  
the Netherlands  
T +31 78 6332332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)



**OIML Member State**  
The Netherlands

Number R51/2006-B-NL1-18.03  
Project number 1901553  
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- No. NMI-13200509-01 dated 20 December 2013 that includes 26 pages;
- No. NMI-1901553-01 dated 9 April 2018 that includes 24 pages;
- No. NMI-1901553-02 dated 9 April 2018 that includes 26 pages.

### Characteristics of the automatic catchweighing instrument

Destined to be used as		Checkweigher, weight/price labeller	
Accuracy class		XIII(x) actual accuracy class determined at the time of putting into use Y(a)	
Maximum capacity		$3 \text{ kg} \leq \text{Max} \leq 15 \text{ kg}$	
Minimum capacity		$\text{Min} \geq 50 \text{ g}$	
Verification scale interval		$e \geq 1 \text{ g}$	
Weighing range		Single interval	
Maximum number of scale intervals (single interval)		$n \leq 3000$ divisions	
Tare		$T \leq -\text{Max}$	
Maximum load transport system speed		29 m/min.	
Maximum rate of operation		65 packages per minute	
Electromagnetic environment class		E1	
Climatic environment		temperature range	+5 °C / +40 °C
		humidity	non-condensing
		intended location	closed
Power supply voltage		200 V – 240 V AC 50/60 Hz	
Software identification	Console protected software	Version number	1.0.3.4
		Checksum	0x5A09571F
	PLC protected software	Version number	1.0.1.3
		Checksum	0x3A19EB51