

**OIML Member State**  
The Netherlands

Number R51/2006-B-NL1-18.05  
Project number 2209642  
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant and Manufacturer	VWS Nederland B.V. Hoogschaijksestraat 25 5374 EC Schaijk The Netherlands
Identification of the certified type	An <b>Automatic catchweighing instrument</b> Type : Octopus Max
Characteristics	See next page

This OIML Certificate is issued under scheme B.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R 51** - Edition 2006 (E) for accuracy class XIII(x), Y(a)

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.

This Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**  
17 December 2018



C. Oosterman  
Head Certification Board

NMi Certin B.V.  
Hugo de Grootplein 1  
3314 EG Dordrecht  
the Netherlands  
T +31 78 6332332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)



**OIML Member State**  
The Netherlands

Number R51/2006-B-NL1-18.05  
Project number 2209642  
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- No. NMI-13200509-01 dated 20 December 2013 that includes 26 pages;
- No. NMI-1901553-01 dated 9 April 2018 that includes 24 pages;
- No. NMI-1901553-02 dated 9 April 2018 that includes 26 pages;
- No. NMI-2209642-01 dated 17 December 2018 that includes 16 pages.

### Characteristics of the automatic catchweighing instrument

Version	FIT7 load cell		AD103C ADPD with BM11 load cell		
Destined to be used as	Checkweigher, weight/price labeller				
Accuracy class	XIII(x) actual accuracy class determined at the time of putting into use Y(a)				
Maximum capacity	3 kg ≤ Max ≤ 15 kg	30 kg ≤ Max ≤ 150 kg			
Minimum capacity	Min ≥ 50 g	Min ≥ 1250 g			
Verification scale interval	e ≥ 1 g	e ≥ 10 g			
Weighing range	Single interval				
Maximum number of scale intervals (single interval)	n ≤ 3000 divisions				
Tare	T ≤ -Max				
Maximum load transport system speed	29 m/min.	40 m/min.			
Maximum rate of operation	65 packages per minute	60 packages per minute			
Electromagnetic environment class	E1				
Climatic environment	temperature range	+5 °C / +40 °C			
	humidity	non-condensing			
	intended location	closed			
Power supply voltage	200 V – 240 V AC 50/60 Hz				
Software identification	Console protected software	Version number	1.0.3.4		
		Checksum	0x5A09571F		
	PLC protected software	Version number	1.0.1.3 or 1.0.1.4	1.0.1.4	
		Checksum	0x3A19EB51 or 0x75587D96	0x75587D96	