



OIML Certificate

OIML Member State
The Netherlands

Number R51/2006-A-NL1-20.01
Project number 2406212
Page 1 of 2

Issuing authority NMI Certin B.V.
Person responsible: F. van Booma – de Smit

Applicant and Manufacturer Società Cooperativa Bilanciali a.r.l.
Via S. Ferrari, 16
41011 Campogalliano, Modena
Italy

Identification of the certified type An **Automatic catchweighing instrument**
Type : GALAXI-LT,
GALAXI-HS

Characteristics See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 51 - Edition 2006 (E) for accuracy class XIII(1), Y(a)

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMI Certin B.V., OIML Issuing Authority NL1**
23 January 2020

Certification Board

NMI Certin B.V.
Thijssseweg 11
2629 JA Delft
The Netherlands
T +31 88 6362332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMI Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.



The conformity was established by the results of tests and examinations provided in the associated OIML Reports:

- No. NMO-P02312 dated 29 June 2018 that includes 14 pages;
- No. NMO-TR761 dated 29 June 2018 that includes 61 pages;
- No. NMI-2406212-01 dated 23 January 2020 that includes 23 pages.

Characteristics of the automatic catchweighing instrument

Operation type and model of instrument	Static weighing GALAXI-LT/HS	Dynamic weighing GALAXI-HS
Destined to be used as	checkweigher, weight labeller, weight/price labeller	
Accuracy class	XIII(1) actual accuracy class determined at the time of putting into use Y(a)	
Maximum capacity	Max ≤ 15 kg	Max ≤ 9 kg
Minimum capacity	Min ≥ 20 e	Min ≥ 20 g (for category Y) Min ≥ 50 g (for category X)
Verification scale interval	e ≥ 1 g	
Weighing ranges	Single interval Multi-interval	
Maximum number of scale intervals (per partial weighing range)	n ≤ 6000 divisions	n ≤ 4500 divisions
Maximum number of partial weighing ranges	2	
Maximum load transport system speed	70 m/min	70 m/min (m ≤ 1 kg) 54 m/min (1 kg < m ≤ 8 kg) 27 m/min (8 kg < m ≤ 9 kg)
Electromagnetic environment class	E2	
Climatic environment	temperature range	0 °C / +40 °C
	humidity	non-condensing
	intended location	Closed
Power supply voltage	220 - 240 V AC 50 Hz	
Software identification	A/D converter board software	SCALE: 491064 x.x (x = 0... 99)
	GALAXI board software	LEGAL.OUT: 5.x.x (x = 0... 99)

Software:

- The identification number will be displayed after pressing the key sequence: time/date button – SW Ver.