



OIML CERTIFICATE OF CONFORMITY

To Beijing True-Tec Co., Ltd

This certificate attests the conformity of the type of the
Load Cell PA06 with the requirements
of the Recommendation R60(2000) of the
International Organization of Legal Metrology(OIML).

The issuing authority

Office of OIML Affairs

中国秘书处

April 6, 2006

The OIML member

王秦平

April 6, 2006

Member State
P.R.China

OIML Certificate No.

✓ R60/2000-CN1-06.10

OK

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: Office of OIML Affairs, General Administration of Quality Supervision, Inspection and Quarantine
Address: N0.9 Madiandonglu, Haidian District, Beijing, 100088 China

Applicant

✓ Name: Beijing True-Tec Co., Ltd
Address: , Room B#-303, No.10A, Long Qing Street, BDA, Beijing 100176 China

Manufacturer of the certified type (if the manufacturer is not the applicant)

Name:

Address:

Identification of the certified type

Load Cell

Model designation	PA06									
Accuracy class	C1									
Maximum capacity (E_{max})	3kg	5kg	7kg	10kg	15kg	20kg	30kg	35kg	50kg	75kg
Minimum load cell verification interval (V_{min})	0.7g	1.2g	1.6g	2.5g	3.5g	5g	7g	8g	15g	25g
Minimum dead load (E_{min})	0									
Maximum number of load cell interval (n_{max})	4000									
Apportionment factor (P_c)	0.7									
Safe overload	150% E_{max}									
Operating Temperature range	-10°C --- +40°C									
Humidity condition	CH									

The numbers of the associated test reports and the included pages:

✓ LXfc2005-6003 18 pages, LXfc2005-6004 17 pages

This certificate attests the conformity of the above-mentioned type(represented by the samples identified in the associated test report) with the requirements of the following Recommendation(s) of the International Organization of Legal Metrology (OIML).

R: 60

Edition: 2000

For accuracy class (if applicable): C4

This certificate relates only to the metrological and technical characteristics of the type of the instrument concerned ,as covered by the relevant OIML International Recommendation(s).

This certificate does not bestow any form of legal international approval. The conformity was established by tests described in the associated test reports.

The issuing authority

Office of OIML Affairs



The CIML member

王春平

April 6, 2006

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.