

Member State  
P.R.China

OIML Certificate No.  
R60/2000-CN1-07.04

# OIML CERTIFICATE OF CONFORMITY

## Issuing authority

Name: Office of OIML Affairs, General Administration of Quality Supervision, Inspection and Quarantine

Address: N0.9 Madiandonglu, Haidian District, Beijing, 100088 China

## Applicant

Name: Hottinger Baldwin Measurement (Suzhou) Co., Ltd.

Address: 106 Hengshan Road Suzhou (215009) P.R.China

## Manufacturer of the certified type (if the manufacturer is not the applicant)

Name:

Address:

## Identification of the certified type

### Load Cell

Model designation	MB35D1		MB35C3		
Accuracy class	D		C		
Maximum capacity ( $E_{max}$ )	220kg, 550kg, 1100kg, 1760kg				
Minimum load cell verification interval ( $V_{min}$ )	$E_{max}$	220kg	550kg	1100kg	1760kg
	D	62.7g	156.7g	313.5g	501.6g
	C	22.0g	55.0g	110.0g	176.0g
Minimum dead load ( $E_{min}$ )	0kg				
Maximum number of load cell interval ( $n_{max}$ )	1000		3000		
Apportionment factor ( $P_{LC}$ )	0.7				
Safe overload	150% $E_{max}$				
Operating Temperature range	-10 ~ +40				

The numbers of the associated test reports and the included pages:

2006-WJ-0011 pages 18 , 2006-WJ-0012 pages 17

This certificate attests the conformity of the above-mentioned type(represented by the samples identified in the associated test report) with the requirements of the following Recommendation(s) of the International Organization of Legal Metrology (OIML).

R: 60

Edition: 2000

For accuracy class (if applicable): D1 , C3

This certificate relates only to the metrological and technical characteristics of the type of the instrument concerned ,as covered by the relevant OIML International Recommendation(s).

This certificate does not bestow any form of legal international approval. The conformity was established by tests described in the associated test reports.

The issuing authority



April 10, 2007

The CIML member

April 10, 2007

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.