

Member State
P.R.China



OIML Certificate No.
R60/2000-CN1-09.02
Revision 1

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: Office of OIML Affairs, General Administration of Quality Supervision, Inspection and Quarantine
Address: No.9 Madiandonglu, Haidian District, Beijing, 100088 P.R.China

Applicant

Name: Beijing True-Tec Co., Ltd
Address: 4/F, Bldg. 2, No. 8, Hong Da Bei Lu, BDA, Beijing (100176) P.R.China

Manufacturer of the certified type (if the manufacturer is not the applicant)

Name:
Address:

Identification of the certified type

Load Cell

Model designation	PA 28
Accuracy class	C3
Maximum capacity (E_{max})	50kg, 100kg, 125kg, 150kg, 200kg, 250kg
Minimum load cell verification interval (V_{min})	Y=5000
Minimum dead load (E_{min})	0
Maximum number of load cell interval (n_{max})	3000
Apportionment factor (P_{LC})	0.7
Safe overload	150% E_{max}
Humidity classification	NH

The numbers of the associated test reports and the included pages:
LSfc2008-6008 17 pages

This certificate attests the conformity of the above-mentioned type(represented by the samples identified in the associated test report) with the requirements of the following Recommendation(s) of the International Organization of Legal Metrology (OIML).

R: 60

Edition: 2000

For accuracy class (if applicable): C3

This certificate relates only to the metrological and technical characteristics of the type of the instrument concerned ,as covered by the relevant OIML International Recommendation(s).

This certificate does not bestow any form of legal international approval.

The conformity was established by tests described in the associated test reports.

This Revision replaces the preceding certificate.

The issuing authority
Office of OIML Affairs

中国秘书处

August 30, 2010

The CIML member

蒲长城

August 30, 2010

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.