

Member State
P.R.China



OIML Certificate No.
R60/2000-CN1-10.01

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: Office of OIML Affairs, General Administration of Quality Supervision, Inspection and Quarantine

Address: N0.9 Madiandonglu, Haidian District, Beijing, 100088 China

Applicant

Name: Hottinger Baldwin Measurement (Suzhou) Co., Ltd.

Address: 106 Hengshan Road Suzhou (215009) P.R.China

Manufacturer of the certified type (if the manufacturer is not the applicant)

Name:

Address:

Identification of the certified type

Load Cell

Model designation	MB35
Maximum capacity (E_{max})	220kg, 550kg, 1100kg, 1760kg
Minimum dead load (E_{min})	0 kg
Accuracy class	C3
Maximum number of load cell interval (n_{max})	3000
Minimum load cell verification interval (v_{min})	0.01% of E_{max}
Apportionment factor (P_{LC})	0.7
Humidity classification	CH
Sensitivity (mV/V)	2

The numbers of the associated test reports and the included pages:

LSfc2010-6001 18 pages , LSfc2010-6002 17 pages,

This certificate attests the conformity of the above-mentioned type(represented by the samples identified in the associated test report) with the requirements of the following Recommendation(s) of the International Organization of Legal Metrology (OIML).

R: 60

Edition: 2000

For accuracy class (if applicable): C3

This certificate relates only to the metrological and technical characteristics of the type of the instrument concerned ,as covered by the relevant OIML International Recommendation(s).

This certificate does not bestow any form of legal international approval. The conformity was established by tests described in the associated test reports.

The issuing authority
Office of OIML Affairs



The CIML member

A handwritten signature in black ink, consisting of three stylized Chinese characters: '蒲长城' (Pu Changcheng).

June 10, 2010

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.