

Member State
P.R.China



OIML Certificate No.
R60/2000-CN1-10.03

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: Office of OIML Affairs, General Administration of Quality Supervision, Inspection and Quarantine

Address: N0.9 Madiandonglu, Haidian District, Beijing, 100088 China

Applicant

Name: Minebea Co., Ltd

Address: 1-1-1, Katase, Fujisawa-shi, Kanagawaken

Manufacturer of the certified type (if the manufacturer is not the applicant)

Name:

Address:

Identification of the certified type

Load Cell

Model designation	CB005
Maximum capacity (E_{max})	60kg, 100kg, 150kg, 250kg, 300kg
Minimum dead load (E_{min})	0 kg
Accuracy class	C4
Maximum number of load cell interval (n_{max})	4000
Minimum load cell verification interval (v_{min})	6g, 10g, 15g, 25g, 30g
Apportionment factor (P_{LC})	0.7
Humidity classification	CH
Sensitivity (mV/V)	2.2
Safe load limit	150% E_{max}
Input resistance (Ω)	400~450
Output resistance (Ω)	345~355

This certificate attests the conformity of the above-mentioned type(represented by the samples identified in the associated test report) with the requirements of the following Recommendation(s) of the International Organization of Legal Metrology (OIML).

R: 60

Edition: 2000

For accuracy class (if applicable): C4

This certificate relates only to the metrological and technical characteristics of the type of the instrument concerned ,as covered by the relevant OIML International Recommendation(s).


This certificate does not bestow any form of legal international approval. The conformity was established by tests described in the associated test reports.

The issuing authority

Office of OIML Affairs
中国秘书处

December 27, 2010

The CIML member



December 27, 2010

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.