

**Member State of OIML  
United Kingdom of Great Britain  
and Northern Ireland**

**OIML Certificate No  
R60/2000-GB1-09.11**

Issuing authority

Name: **National Weights and Measures Laboratory  
(Part of the National Measurement Office)**  
Address: **Stanton Avenue  
Teddington  
Middlesex, TW11 0JZ  
United Kingdom**

Person responsible: **Paul Dixon – Product Certification Manager**

Applicant

Name: **Fotiadis Panagiotis**  
Address: **Industrial Area Sindos  
P.Box. 1217  
570 22 Thessaloniki  
Greece**

Manufacturer of the certified pattern is:

**The applicant**

Identification of the certified pattern:

**Stainless steel shear beam load cell**

Model Designation	FAP-04					
Maximum capacity, $E_{\max}$ (kg)	500	750	1000	1500	2000	2500
Accuracy class	C3					
Maximum number of load cell intervals, $n_{\max}$	3000					
Minimum verification interval, $V_{\min}$	$E_{\max} / 10000$					
Apportionment factor; $p_{LC}$	0.70					

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report) with the requirements of the following Recommendation of the International Organisation of Legal Metrology -OIML):

**R 60 Metrological regulation for load cells Edition: 2000 (E) for accuracy class : C3**

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

# OIML Certificate No R60/2000-GB1-09.11

This certificate does not bestow any form of legal international approval.

The conformity was established by tests described in the associated:

Test report: CEM-IYO-04/0156-5.1 having 18 pages (issued by CEM)

Issuing authority



Mr P Dixon  
for NWML

CIML member



Mr P Mason

Date 27 November 2009

Ref: T1136/0050

### Essential technical data

<i>Model designation</i>	<i>Designation</i>	<i>Value</i>	<i>Units</i>
Classification		C3	
Additional marking		CH	
Maximum number of load cell verification intervals	$n_{LC}$	3000	
Maximum capacity	$E_{max}$	500, 750, 1000, 1500, 2000, 2500	kg
Minimum dead load, relative	$E_{min}/E_{max}$	0	
Relative $V_{min}$ (ratio to minimum LC verification interval)	$Y = E_{max}/V_{min}$	10000	g
Relative DR (ratio to minimum dead load output return)	$Z = E_{max}/(2*DR)$	12462	kg
Rated output		2	mV/V
Maximum excitation voltage		18	V DC
Input impedance (for strain gauge LCs)	$R_{LC}$	$386 \pm 2\%$	$\Omega$
Temperature rating		-10/+40	$^{\circ}C$
Safe overload, relative	$E_{lim}/E_{max}$	125	% F.S
Cable length (maximum)		4	m
Additional characteristics		6 wire (plus screen)	

### Certificate History

ISSUE NO.	DATE	DESCRIPTION
R60/2000-GB1-09.11	27 November 2009	Type approval first issued
-	-	No revisions have been issued

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.