



Member State of OIML United Kingdom of Great Britain and Northern Ireland OIML Certificate No R60/2000-GB1-12.01

## **OIML CERTIFICATE OF CONFORMITY**

Issuing authority:	National Measurement Office	
Person responsible:	Paul Dixon – Product Certification Manager	
Applicant:	Vehicle Weighing Solutions Ltd Southview Park Marsack Street Caversham Reading, RG4 5AF United Kingdom	
Manufacturer:	The applicant	
Identification of the certified pattern:	AW685 stainless steel compression load cell	

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report) with the requirements of the following Recommendation of the International Organisation of Legal Metrology (OIML):

## OIML R 60 - Edition 2000(E) for accuracy class: C3

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

This certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the certificates reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.

Issue Date: Reference No:

02 April 2012 TS13/0014

Signatory: P R Dixon



National Measurement Office | Stanton Avenue | Teddington | TW11 0JZ | United Kingdom Tel +44 (0)20 8943 7272 | Fax +44 (0)20 8943 7270 | Web www.bis.gov.uk/nmo The conformity was established by tests described in the associated test report SN: 1260 which includes 23 pages.

Model designation	Designation	Value	Units
Classification		C3	
Additional marking		СН	
Maximum number of load cell verification intervals	n <sub>LC</sub>	3000	
Maximum capacity	E <sub>max</sub>	1000, 2000, 5000	kg
Minimum dead load, relative	E <sub>min</sub> /E <sub>max</sub>	0	kg
Relative V <sub>min</sub> (ratio to minimum LC verification interval)	$Y = E_{max}/V_{min}$	E <sub>max</sub> /3030	kg
Relative DR (ratio to minimum dead load output return)	$Z = E_{max}/(2*DR)$	4552	
Rated output		0.95 +/-0.05	mV/V
Maximum excitation voltage		+ 15	V DC
Input impedance (for strain gauge LCs)	R <sub>LC</sub>	380 +/-30	Ω
Temperature rating		-10/+40	°C
Safe overload, relative	E <sub>lim</sub> /E <sub>max</sub>	150	% F.S
Fraction	P <sub>LC</sub>	0.7	
Cable length		6	m
Additional characteristics		6 wire	

## **Characteristics of the Load Cell:**

## **Certificate History**

Issue №.	Date	Description
R60/2000-GB1-12.01	02 April 2012	Type approval first issued
-	-	