



Member State of OIML
United Kingdom of Great Britain
and Northern Ireland

OIML Certificate № R60/2000-GB1-12.07 Revision 1

## OIML CERTIFICATE OF CONFORMITY

Issuing authority: National Measurement Office

Person responsible: Paul Dixon – Product Certification Manager

Applicant: Flintec GmbH

Bemannsbruch 9

DE-74909 Meckesheim

Germany

Manufacturer: The applicant

Identification of the

certified pattern: SB6 stainless steel load cell

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report) with the requirements of the following Recommendation of the International Organisation of Legal Metrology (OIML):

## OIML R 60 - Edition 2000(E) for accuracy class: C3 MI6

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

This certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the certificates reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.

Issue Date: 08 March 2013 Reference №: TS13/0016

Signatory: P R Dixon

National Measurement Office | Stanton Avenue | Teddington | TW11 0JZ | United Kingdom Tel +44 (0)20 8943 7272 | Fax +44 (0)20 8943 7270 | Web www.bis.gov.uk/nmo





The conformity was established by tests described in the associated test report SN: 1238 issued by NMO.

## **Characteristics of the Load Cell**

Model designation	Designation	Value	Units
Classification		C3 MI6	
Additional marking		-	
Maximum number of load cell verification intervals	n <sub>LC</sub>	3000	
Maximum capacity	E <sub>max</sub>	50, 100, 150, 200 & 250	kg
		0.5, 1 & 2	kN
Minimum dead load, relative	E <sub>min</sub> /E <sub>max</sub>	0	%
Relative V <sub>min</sub> (ratio to minimum LC verification interval)	$Y = E_{max}/V_{min}$	20400	
Relative DR (ratio to minimum dead load output return)	$Z = E_{\text{max}}/(2*DR)$	6270	
Rated output		$2.0 \pm 0.02$	mV/V
Maximum excitation voltage		15	V DC
Input impedance (for strain gauge LCs)	R <sub>LC</sub>	1100 ± 50	Ω
Temperature rating		-10/+40	°C
Safe overload, relative	E <sub>lim</sub> /E <sub>max</sub>	200	%
Fraction	P <sub>LC</sub>	0.7	
Cable length		3	m
Additional characteristics		4 or 6 wire	

## **Certificate History**

Issue №.	Date	Description
R60/2000-GB1-12.07	28 November 2012	Certificate first issued
R60/2000-GB1-12.07 Revision 1	08 March 2013	Additional capacities added.