



#### **OIML Member State**

United Kingdom of Great Britain and Northern Ireland

OIML Certificate No. R60/2000-A-GB1-18.03

### **OIML CERTIFICATE ISSUED UNDER SCHEME A**

OIML Issuing Authority NMO

Stanton Avenue Teddington TW11 0JZ United Kingdom

January III.

Person responsible: Mannie Panesar – Head of Technical Services

Applicant CARDINAL SCALE MANUFACTURING COMPANY

**203 EAST DAUGHERTY STREET** 

WEBB CITY, MISSOURI MO 64870

USA

Manufacturer The applicant

Identification of the AC

certified type (the detailed characteristics are defined in the Descriptive Annex)

This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 60, Edition: 2000

For accuracy class: C3

Issue date: 25 May 2018

The OIML Issuing Authority

**Grégory Glas** 

**Lead Technical Manager** 

For and on behalf of the Head of Technical Services

U K A S
PRODUCT

0135

This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated OIML type evaluation report:

No. P02197 dated 25 May 2018 that includes 3 pages

The technical documentation relating to the identified type is contained in documentation file: No. P02197-D dated 25 May 2018.

## **OIML Certificate History**

Revision No.	Date	Description of the modification		
Revision 0	25 May 2018	Certificate first issued		
-	-	-		

No revisions have been issued.

#### Important note:

Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full.

## **DESCRIPTIVE ANNEX**

# **Characteristics of the Load Cell:**

	Designation			Valu	e		Units
Accuracy Class	3	C3					
Additional marking		CH					
Maximum number of load cell verification intervals	n <sub>LC</sub>	3 000					
Maximum capacity	E <sub>max</sub>	30	40	50	100	150	t
Minimum dead load, relative	E <sub>min</sub> /E <sub>max</sub>	0			%		
Minimum load cell verification interval	V <sub>min</sub>	3	4	5	10	15	kg
Relative v <sub>min</sub> (ratio to minimum load cell verification interval)	$Y = E_{max}/v_{min}$	10 000					
Relative DR (ratio to minimum dead load output return)	$Z = E_{\text{max}}/(2*DR)$	3 000					
Rated output		2.0					mV/V
Maximum excitation voltage		15				V ac/dc	
Input impedance (for strain gauge load cells)	R <sub>LC</sub>	1150 ± 50				Ω	
Temperature rating		-10 / + 40			°C		
Safe overload, relative	$E_{lim}/E_{max}$	200			% F.S		
Apportionment factor	P <sub>LC</sub>	0.7					
Cable length:		≤ 18					m
Additional characteristics:	6 wire						
Transducer material	Stainless steel						
Atmospheric protection	Hermetic Welded						
Output impedance	1005 ± 5					Ω	
Reference excitation voltage	10					V ac/dc	
Cable cross-section	0.25					mm <sup>2</sup>	

## **CERTIFICATE HISTORY**

ISSUE NO.	DATE	DESCRIPTION				
R60/2000-A-GB1-18.03	25 May 2018	Certificate first issued.				
-	-	No revisions have been issued.				