



OIML Member State

United Kingdom of Great Britain and Northern Ireland

OIML Certificate No. R60/2000-A-GB1-20.01

OIML CERTIFICATE ISSUED UNDER SCHEME A

OIML Issuing Authority NMO

Stanton Avenue Teddington TW11 0JZ

United Kingdom

Person responsible: Mannie Panesar – Head of Technical Services

Applicant Ugno-Uralsky Vesovoy Zavod (Ural Sable)

134 Building 6, Mendeleev, Ufa, 450022

Russian Federation

Manufacturer Ugno-Uralsky Vesovoy Zavod (Ural Sable)

134 Building 6, Mendeleev, Ufa, 450022

Russian Federation

Identification of the ST-M-H(K)(P)-B Canister compression

certified type (the detailed characteristics are defined in the Descriptive Annex)

This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 60, Edition: 2000

For accuracy class: C3

Issue date: 13 January 2020

The OIML Issuing Authority

Grégory Glas

Lead Technical Manager

For and on behalf of the Head of Technical Services

This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated OIML type evaluation report:

No. P02686 dated 13 January 2020 that includes 3 pages

The technical documentation relating to the identified type is contained in documentation file: No. P02686-D dated 13 January 2020.

OIML Certificate History

Revision No. Date		Description of the modification
Revision 0	on 0 13 January 2020 Certificate first issued	
-	-	-

No revisions have been issued.

Important note:

Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full.

ST-designation, column type sensor (identification codes):

ST-M-H(K)(P)-B

M - maximum load, t;

H - internal resistance: B;

K - cable entry location: 1, 2;

P - device design from turning the sensor: 1, 2;

B - explosion-proof execution.

DESCRIPTIVE ANNEX

Characteristics of the Load Cell:

	Designation	Value				Units
Accuracy Class		C3				
Additional marking		СН				
Maximum number of load cell verification intervals	n _{LC}	3 000				
Maximum capacity	E _{max}	20	30	50	100	t
Minimum dead load, relative	E _{min} /E _{max}	0			%	
Minimum load cell verification interval	V _{min}	2.5	3.75	6.25	12.5	kg
Relative v _{min} (ratio to minimum load cell verification interval)	$Y = E_{max}/v_{min}$	8 000				
Relative DR (ratio to minimum dead load output return)	$Z = E_{\text{max}}/(2*DR)$	4000				
Rated output		2.0 ± 0.5			mV/V	
Excitation voltage		5 – 10			V dc	
Input impedance (for strain gauge load cells)	R _{LC}	700 ± 15			Ω	
Temperature rating		-10 / + 40			°C	
Safe overload, relative	E _{lim} /E _{max}	150			% F.S	
Apportionment factor	P _{LC}	0.7				
Cable length:		≤ 20			m	
Additional characteristics:	6 wire					
Transducer material	Stainless steel					
Atmospheric protection	Hermetic Welded and IP68					
Output impedance	840 ± 10				Ω	

The above identified Type (represented by the sample(s) identified in the OIML Test Report) have been found to comply with the additional national requirements established by the United States of America (NIST Handbook 44 and NCWM Publication 14), included in the Utilizer Declaration:

- R60 OIML-CS rev. 2, Additional requirements from the United States Accuracy class III L;
- R60 OIML-CS rev. 2, Additional requirements from the United States Marking requirements.

CERTIFICATE HISTORY

ISSUE NO.	DATE	DESCRIPTION			
R60/2000-A-GB1-20.01	13 January 2020	Certificate first issued.			
-	-	No revisions have been issued.			