



Member State of OIML  
Japan



OIML Certificate No.  
R60/2000-JP1-10.04  
Revision1

## OIML CERTIFICATE OF CONFORMITY

### Issuing authority

Name: National Metrology Institute of Japan / National Institute of  
Advanced Industrial Science and Technology (NMIJ / AIST)  
Address: AIST Tsukuba Central 3-9, Tsukuba Ibaraki 305-8563, Japan  
Person responsible: Dr. Tamotsu Nomakuchi, President of AIST

### Applicant

Name: A&D Company, Limited  
Address: 3-23-14, Higashi-Ikebukuro, Toshima-ku, Tokyo 170-0013, Japan

### Manufacturer of the certified pattern

Name: A&D Company, Limited  
Address: 3-23-14, Higashi-Ikebukuro, Toshima-ku, Tokyo 170-0013, Japan

### Identification of the certified pattern:

Compression Load cells  
Type: LCC11T010-KC, LCC11T020-KC, LCC11T030-KC  
Fraction: Pi=0.7  
Temperature range: -10 °C / 40 °C



Member State of OIML  
Japan

OIML Certificate No.  
R60/2000-JP1-10.04  
Revision1

Characteristics:

Model designation			LCC11T010-KC	LCC11T020-KC	LCC11T030-KC
Accuracy class	Class	-	C		
Maximum number of load cell verification intervals	$n_{max}$	-	4000,5000,6000		
Humidity symbol			CH		
Minimum dead load	$E_{min}$	kg	0		
Maximum capacity	$E_{max}$	t	10	20	30
Safe load limit	$E_{lim}$	t	15	30	45
Minimum verification interval	$v_{min}$	kg	1	2	3
Apportionment factor	$p_{LC}$		0.7		
Ratio of minimum LC Verification interval $Y=E_{max} / v_{min}$	$Y$	-	10000		
Ratio of minimum dead load output return $Z=E_{max}/(2*DR)$	$Z$	-	6000 in the case of $n_{max}=6000$		
Rated output		mV/V	1.96		
Maximum excitation voltage		V DC	15		
Input impedance	$R_{LC}$	$\Omega$	800		
Cable length		m	12		

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report(s) with the requirements of the following Recommendation of the International Organization of Legal Metrology - OIML):

R60, edition 2000 (E)  
For accuracy class C

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

This certificate does not bestow any form of legal international approval.

The conformity was established by tests described in the associated test report no. 11-08/R60:2000, that includes 27 pages.



Member State of OIML  
Japan

OIML Certificate No.  
R60/2000-JP1-10.04  
Revision1

The Issuing Authority  
NMIJ/AIST

The CIML member



Dr. T. Nomakuchi  
President of AIST  
2011-08-09

Dr. Y. Miki  
2011-08-09

Important note: Apart from the mention of certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or the associated test report is not permitted, though they may be reproduced in full.