



Member State of OIML
Japan



OIML Certificate No
R60/2000-JP1-10.16

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: National Metrology Institute of Japan / National Institute of
Advanced Industrial Science and Technology (NMIJ / AIST)
Address: AIST Tsukuba Central 3-9, Tsukuba Ibaraki 305-8563, Japan
Person responsible: Dr. Tamotsu Nomakuchi, President of AIST

Applicant

Name: MINEBEA CO., LTD.
Address: 1-1-1, Katase, Fujisawa-shi, Kanagawa-ken, 251-8531, Japan

Manufacturer of the certified pattern

Name: MINEBEA CO., LTD.
Address: 1-1-1, Katase, Fujisawa-shi, Kanagawa-ken, 251-8531, Japan

Identification of the certified pattern:

Compression load cell
Type: DC002-10T, DC002-20T, DC002-25T, DC002-30T, DC002-40T
Fraction: $\text{Pi}=0.7$
Temperature range: $-10\text{ }^{\circ}\text{C} / 40\text{ }^{\circ}\text{C}$



Member State of OIML
Japan



OIML Certificate No
R60/2000-JP1-10.16

Characteristics:

Model designation			DC002-10T	DC002-20T	DC002-25T	DC002-30T	DC002-40T
Accuracy class	Class	-	C				
Maximum number of load cell verification intervals	n_{max}	-	5000				
Humidity symbol			CH				
Minimum dead load	E_{min}	kg	0				
Maximum capacity	E_{max}	kg	10000	20000	25000	30000	40000
Safe load limit	E_{lim}	kg	15000	30000	37500	45000	60000
Minimum verification interval	v_{min}	kg	0.5	1	1.25	1.5	2
Apportionment factor	p_{LC}		0.7				
Ratio of minimum LC Verification interval $Y = E_{max} / v_{min}$	Y	-	20000				
Ratio of minimum dead load output return $Z = E_{max} / (2 * DR)$	Z	-	5000				
Excitation voltage		V DC	7.5 ~ 15				
Cable length		m	12				15

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report(s) with the requirements of the following Recommendation of the International Organization of Legal Metrology - OIML):

R60, edition 2000 (E)
For accuracy class C

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

This certificate does not bestow any form of legal international approval.

The conformity was established by tests described in the associated test report no. 10-19/R60:2000, that includes 33 pages.



Member State of OIML
Japan



OIML Certificate No
R60/2000-JP1-10.16

The Issuing Authority
NMIJ/AIST

The OIML member



Dr. T. Nomakuchi
President of AIST
2010-08-23

Dr. Y. Miki
2010-08-23

Important note: Apart from the mention of certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.