



Member State of OIML  
Japan

OIML Certificate No  
R60/2000-JP1-10.18

## OIML CERTIFICATE OF CONFORMITY

### Issuing authority

Name: National Metrology Institute of Japan / National Institute of  
Advanced Industrial Science and Technology (NMIJ / AIST)  
Address: AIST Tsukuba Central 3-9, Tsukuba Ibaraki 305-8563, Japan  
Person responsible: Dr. Tamotsu Nomakuchi, President of AIST

### Applicant

Name: YAMATO SCALE CO., LTD.  
Address: 5-22, Saenba-cho, Akashi, Hyogo, 673-8688, Japan

### Manufacturer of the certified pattern

Name: YAMATO SCALE CO., LTD.  
Address: 5-22, Saenba-cho, Akashi, Hyogo, 673-8688, Japan

### Identification of the certified pattern:

Beam (bending) load cell  
Type: UB2-100, UB2-200, UB2-300, QUB2-100, QUB2-200, QUB2-300  
Fraction:  $\pi=0.7$   
Temperature range:  $-10\text{ }^{\circ}\text{C} / 40\text{ }^{\circ}\text{C}$



Member State of OIML  
Japan



OIML Certificate No  
R60/2000-JP1-10.18

Characteristics:

Model designation			UB2-100 QUB2-100	UB2-200 QUB2-200	UB2-300 QUB2-300
Accuracy class	Class	-	C		
Maximum number of load cell verification intervals	$n_{max}$	-	5000		
Humidity symbol			CH		
Minimum dead load	$E_{min}$	kg	0		
Maximum capacity	$E_{max}$	kg	100	200	300
Safe load limit	$E_{lim}$	kg	150	300	450
Minimum verification interval	$v_{min}$	g	10	20	30
Apportionment factor	$p_{LC}$		0.7		
Ratio of minimum LC Verification interval $Y = E_{max} / v_{min}$	$Y$	-	10000		
Ratio of minimum dead load output return $Z = E_{max} / (2 * DR)$	$Z$	-	5000		
Rated output		mV/V	2.0		
Excitation voltage		V AC/DC	4 ~ 15		
Input impedance	$R_{LC}$	$\Omega$	350 ± 10		
Cable length (maximum)		m	5		

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report(s) with the requirements of the following Recommendation of the International Organization of Legal Metrology - OIML):

R60, edition 2000 (E)  
For accuracy class C

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

This certificate does not bestow any form of legal international approval.

The conformity was established by tests described in the associated test report no. 10-22/R60:2000, that includes 19 pages.



Member State of OIML  
Japan



OIML Certificate No  
R60/2000-JP1-10.18

The Issuing Authority  
NMIJ/AIST



Dr. T. Nomakuchi  
President of AIST  
2010-09-22

The OIML member

Dr. Y. Miki  
2010-09-22

Important note: Apart from the mention of certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.