



Member State of OIML  
Japan



OIML Certificate No.  
R60/2000-JP1-12.04

## OIML CERTIFICATE OF CONFORMITY

### Issuing authority

Name: National Metrology Institute of Japan / National Institute of  
Advanced Industrial Science and Technology (NMIJ / AIST)  
Address: AIST Tsukuba Central 3-9, Tsukuba Ibaraki 305-8563, Japan  
Person responsible: Dr. Tamotsu Nomakuchi, President of AIST

### Applicant

Name: MINEBEA CO., LTD.  
Address: 1-1-1, Katase, Fujisawa-shi, Kanagawa-ken, 251-8531, Japan

### Manufacturer of the certified pattern

Name: MINEBEA CO., LTD.  
Address: 1-1-1, Katase, Fujisawa-shi, Kanagawa-ken, 251-8531, Japan

### Identification of the certified pattern:

Compression load cell  
Type: CC010-10T, CC010-20T, CC010-30T, CC010-50T  
Fraction:  $\pi=0.7$   
Temperature range:  $-10\text{ }^{\circ}\text{C} / 40\text{ }^{\circ}\text{C}$



Member State of OIML  
Japan



OIML Certificate No.  
R60/2000-JP1-12.04

Characteristics:

Model designation			CC010-xx T, where xx equivalent to the $E_{max}/1000$
Accuracy class	Class	-	C
Maximum number of load cell verification intervals	$n_{max}$	-	3000
Humidity symbol			CH
Minimum dead load	$E_{min}$	kg	0
Maximum capacity	$E_{max}$	kg	10000,20000,30000,50000
Safe load limit	$E_{lim}$	kg	$1.5 * E_{max}$
Minimum verification interval	$v_{min}$	kg	$E_{max}/10000$
Apportionment factor	$p_{LC}$		0.7
Ratio of minimum LC Verification interval $Y = E_{max} / v_{min}$	Y	-	10000
Ratio of minimum dead load output return $Z = E_{max} / (2 * DR)$	Z	-	3000
Rated output		mV/V	2.0
excitation voltage		V DC	5~15
Input impedance	$R_{LC}$	$\Omega$	$700 \pm 7$
Cable detail		-	10m 4wire

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report(s) with the requirements of the following Recommendation of the International Organization of Legal Metrology - OIML):

R60, edition 2000 (E)  
For accuracy class C

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

This certificate does not bestow any form of legal international approval.

The conformity was established by tests described in the associated test report no. 12-10/R60:2000, that includes 19 pages.



Member State of OIML  
Japan



OIML Certificate No.  
R60/2000-JP1-12.04

The Issuing Authority  
NMIJ/AIST

The OIML member



Dr. T. Nomakuchi  
President of AIST  
2012-05-21

Dr. Y. Miki  
2012-05-21

Important note: Apart from the mention of certificate's reference number and the name of the OIML Member State in which the certificate is issued, partial quotation of the certificate and of the associated Test Report is not permitted, although either may be reproduced in full.