

**OIML CERTIFICATE OF CONFORMITY**

## Issuing authority

Name: NMI Certin B.V.

Address: Hugo de Grootplein 1  
3314 EG Dordrecht  
The Netherlands

## Applicant

Name: Flintec GmbH.

Address: Bemannsbruch 9  
74909 Meckesheim  
Germany

## Manufacturer of the certified type

Name: Flintec GmbH.

Address: Bemannsbruch 9  
74909 Meckesheim  
Germany

## Identification of certified type

Load cell, with a digital output  
Type : RC3D  
Fraction :  $P_i = 0.8$   
Temperature range  $-10\text{ °C} / +40\text{ °C}$ 

Maximum capacity ( $E_{max}$ )	25000 kg up to and including 50000 kg		
Accuracy Class	C		
Maximum number of load cell intervals (n)	1000	3000	4000
Ratio of minimum LC Verification interval $Y = E_{max} / V_{min}$	5000	15000	15000

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the test certificate and the description with number TC6586 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**R60**  
Edition 2000 (E)  
for accuracy class C

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Report:  
N° R60/2000-NL1-04.16 that includes 53 pages.

This revision replaces the earlier versions, except for its test report.

The Issuing Authority NL1  
NMI Certim, 21 September 2009



C. Oosterman  
Head Certification Board

\*  
\* \*

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.