

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMI Certin B.V.
Address: Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
Person responsible: Ing. C. Oosterman

Applicant

Name: SCAIME S.A.S.
Address: Z.I de Juvigny – B.P. 501 Juvigny
F-74105 Annemasse Cedex
France

Manufacturer of the certified type

Name: SCAIME S.A.S.
Address: Z.I de Juvigny – B.P. 501 Juvigny
F-74105 Annemasse Cedex
France

Identification of certified type

Compression load cell
Family of type : CA40X
Fraction : $P_i = 0.7$
Temperature range : $-10\text{ °C} / +40\text{ °C}$

For further characteristics, see next page where:

E_{\max} = Maximum capacity

E_{\min} = Minimum capacity

n = Maximum number of load cell intervals

$Y = E_{\max} / V_{\min}$ = Ratio of minimum LC Verification interval

$Z = E_{\max} / (2 * DR)$ = Ratio of minimum dead load output return

NMI Certin B.V.

Hugo de Grootplein 1, 3314 EG Dordrecht
P.O. Box 394, 3300 AJ Dordrecht, NL
phone +31 78 6332332
fax +31 78 6332309
certin@nmi.nl
www.nmi.nl

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMI B.V. (see "Regulation objection and appeal against decisions of NMI B.V.")

NMI Certin B.V., chamber o.c. nr. 27.233.418

This document is issued under the provision that no responsibility is accepted and that the applicant gives warranty for each responsibility against third parties.

The notification of NMI Certin as Issuing Authority can be verified at www.oiml.org.



Nederlands Meetinstituut


OIML Member state
The Netherlands

OIML Certificate N° R60/2000-NL1-05.22 revision 1

Project number 510226

Page 2 of 3

Type	E_{min} [kg]	E_{max} [kg]	V_{min} [kg]	Y	n	Z
CA40X5t C4 CH 10e	100	5100	0.5	10200	4000	4080
CA40X5t C3 CH 10e	100	5100	0.5	10200	3000	3060
CA40X5t C3 CH 5e	100	5100	1	5100	3000	3060
CA40X5t C2 CH 5e FP	100	5100	1	5100	2000	2040
CA40X10t C4 CH 10e	200	10200	1	10200	4000	4080
CA40X10t C3 CH 10e	200	10200	1	10200	3000	3060
CA40X10t C3 CH 5e	200	10200	2	5100	3000	3060
CA40X10t C2 CH 5e FP	200	10200	2	5100	2000	2040
CA40X20t C4 CH 10e	500	20500	2	10250	4000	4100
CA40X20t C3 CH 10e	500	20500	2	10250	3000	3075
CA40X20t C3 CH 5e	500	20500	4	5125	3000	3075
CA40X20t C2 CH 5e FP	500	20500	4	5125	2000	2050
CA40X30t C4 CH 10e	500	30500	3	10170	4000	4070
CA40X30t C3 CH 10e	500	30500	3	10170	3000	3050
CA40X30t C3 CH 5e	500	30500	6	5085	3000	3050
CA40X30t C2 CH 5e FP	500	30500	6	5085	2000	2035
CA40X40t C4 CH 10e	1000	41000	4	10250	4000	4100
CA40X40t C3 CH 10e	1000	41000	4	10250	3000	3075
CA40X40t C3 CH 5e	1000	41000	8	5125	3000	3075
CA40X60t C4 CH 10e	1000	61000	6	10170	4000	4070
CA40X60t C3 CH 10e	1000	61000	6	10170	3000	3050
CA40X60t C3 CH 5e	1000	61000	12	5085	3000	3050
CA40X100t C4 CH 10e	1000	101000	10	10100	4000	4040
CA40X100t C3 CH 10e	1000	101000	10	10100	3000	3030
CA40X100t C3 CH 5e	1000	101000	20	5050	3000	3030



Nederlands Meetinstituut

OIML Member state
The Netherlands

OIML Certificate N° R60/2000-NL1-05.22 revision 1

Project number 510226

Page 3 of 3

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the test certificate and the description with number TC5863 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R60
Edition 2000 (E)
for accuracy class C

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Reports:

N° R60/2000-NL-01.08A, that includes 37 pages;

N° R60/2000-NL-01.08B, that includes 40 pages;

N° R60/2000-NL1-05.22, that includes 13 pages.

This revision replaces the earlier version.

The Issuing Authority
Ing. C. Oosterman
Manager Product Certification

20 January 2006

*

**

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.