

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMi Certin B.V.
Address: Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
Person responsible: Ing. C. Oosterman

Applicant

Name: Vishay Tedea-Huntleigh International Ltd.
Address: 5 Hazoran Street
New Industrial Zone Netanya
42506 ISRAEL


Manufacturer of the certified type

Name: Vishay-Transducers
Address: 5 Hazoran Street
New Industrial Zone Netanya
42506 ISRAEL

Identification of certified type

Type : 1004
Fraction : $P_i = 0.7$
Temperature range +5 °C / +40 °C

Maximum capacity (E_{max})	0.3 kg up to and including 3 kg
Accuracy Class	C
Maximum number of load cell intervals (n)	3000
Ratio of minimum LC Verification interval $Y = E_{max} / V_{min}$	5000



Nederlands Meetinstituut

OIML Member state
The Netherlands

OIML Certificate N° R60/2000-NL1-05.23

Project number 510116

Page 2 of 2

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the test certificate and the description with number TC6869 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R60
Edition 2000 (E)
for accuracy class C

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Reports:

N° R60/2000-NL1-05.23A that includes 40 pages;

N° R60/2000-NL1-05.23B that includes 37 pages.

The Issuing Authority
Ing. C. Oosterman
Manager Product Certification


27 December 2005

*

**

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.