



# OIML Certificate of Conformity

**OIML Member State**  
The Netherlands

Number R60/2000-NL1-10.28  
Project number 10200374  
Page 1 of 2

Issuing authority NMI Certin B.V.  
Person responsible: C. Oosterman

Applicant Teraoka Seiko Co. Ltd.  
13-12 Kugahara, 5-Chome, Ohta-Ku, Tokyo 146-8580, Japan

Manufacturer Teraoka Seiko Co. Ltd.  
13-12 Kugahara, 5-Chome, Ohta-Ku, Tokyo 146-8580, Japan

Identification of the certified type **A single point Load Cell**  
Type : PN..., PMB...

Characteristics See next page


This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R60** - Edition 2000 (E) for accuracy class C

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.  
This Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMI Certin B.V., OIML Issuing Authority NL1**  
10 January 2011

  
C. Oosterman  
Head Certification Board

NMI Certin B.V.  
Hugo de Grootplein 1  
3314 EG Dordrecht  
the Netherlands  
T +31 78 6332332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMI Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMI (see [www.nmi.nl](http://www.nmi.nl)).





# OIML Certificate of Conformity

**OIML Member State**  
The Netherlands

Number R60/2000-NL1-10.28  
Project number 10200374  
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- No. R60/2000-NL1-10.28A that includes 43 pages;
- No. R60/2000-NL1-10.28B that includes 43 pages;
- No. R60/2000-NL1-10.28C that includes 38 pages.

**Characteristics of the Load Cell:**

Maximum capacity ( $E_{max}$ )	9 kg	22.5 kg up to and including 90 kg	225 kg up to and including 450 kg
Fraction $P_{LC}$	0.7		
Humidity classification	SH		
Temperature range	0 °C / +40 °C	-10 °C / +40 °C	
Accuracy class	C		
Maximum number of load cell verification intervals ( $n_{max}$ )	3000	7500	3000
Ratio of minimum LC verification interval $Y = E_{max} / v_{min}$	4500	11250	4500
Ratio of minimum dead load output return $Z = E_{max} / 2 * DR$	5000	7500	4500