



OIML Certificate of Conformity

OIML Member State
The Netherlands

Number R60/2000-NL1-11.07
Project number 11200201
Page 1 of 3

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant	Tedea-Huntleigh or Vishay Precision 2 Haofan St., Holon, Israel
Manufacturer	Tedea-Huntleigh or Vishay Precision 2 Haofan St., Holon, Israel
Identification of the certified type	A bending beam Load Cell Type : 1242
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R60 - Edition 2000 (E) for accuracy class C

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
13 April 2011

C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).





OIML Certificate of Conformity

OIML Member State
The Netherlands

Number R60/2000-NL1-11.07
Project number 11200201
Page 2 of 3

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- No. R60/1991-NL-00.05 dated 15 April 1999 that includes 38 pages;
- No. NM-11200201-01 dated 13 April 2011 that includes 29 pages.

Characteristics of the Load Cell:

Type	1242					
Fraction P_i	0,7					
Maximum capacity E_{max}	50 kg up to and including 250 kg					
Humidity Class	CH					
Temperature range	-10 °C / +40 °C					
Accuracy Class	C					
Maximum number of load cell verification intervals n_{max}	1000	2000	3000	4000	5000	6000
Ratio of minimum LC verification interval $Y = E_{max} / v_{min}$	20000					
Ratio of minimum dead load output return $Z = E_{max} / 2 * DR$	15000					



OIML Certificate of Conformity

OIML Member State
The Netherlands

Number R60/2000-NL1-11.07
Project number 11200201
Page 3 of 3

Type	1242					
Fraction P_i	0,7					
Maximum capacity E_{max}	300 up to and including 500 kg					
Humidity Class	CH					
Temperature range	-10 °C / +40 °C					
Accuracy Class	C					
Maximum number of load cell verification intervals n_{max}	1000	2000	3000	4000	5000	6000
Ratio of minimum LC verification interval $Y = E_{max} / v_{min}$	7000					
Ratio of minimum dead load output return $Z = E_{max} / 2 * DR$	8000					

The characteristics for n_{max} and Y and Z can be reduced separately.