



OIML Member State
The Netherlands

OIML Certificate of Conformity

Number R60/2000-NL1-17.58 revision 1
Project number 2845853
Page 1 of 3

Issuing authority

NMi Certin B.V.
Person responsible: M.Ph.D. Schmidt

Applicant and
Manufacturer

Flintec UK Limited
Caxton House, Caxton Place
CF23 8HG Cardiff
United Kingdom

Identification of the
certified type

A **compression load cell**, with strain gauges, equipped with electronics,
Type : RC3D

Characteristics

See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 60 - Edition 2000 (E) for accuracy class C

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

NMi Certin B.V., OIML Issuing Authority NL1
30 November 2021

Certification Board

NMi Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 6362332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.





OIML Member State
The Netherlands

OIML Certificate of Conformity

Number R60/2000-NL1-17.58 revision 1
Project number 2845853
Page 2 of 3

The conformity was established by the results of tests and examinations provided in the associated OIML Test Reports:

- No. R60/2000-NL1-04.16 dated 27 September 2004 that includes 53 pages;
- No. NMI-1901068-01 revision 1 dated 30 November 2021 that includes 89 pages.

Characteristics of the load cell:

Maximum capacity (E_{max})	25 t up to and including 100 t
Minimum dead load	0 kg
Accuracy Class	C
Rated Output	166666 cts if $E_{max} = 25t$ 200000 cts if $E_{max} > 25 t$
Maximum number of load cell intervals (n) ⁽¹⁾	4000
Ratio of minimum LC Verification interval ⁽¹⁾ $Y = E_{max} / V_{min}$	15000
Ratio of minimum dead load output return ⁽¹⁾ $Z = E_{max} / (2 * DR)$	4000
Temperature range	-10 °C / + 40 °C
Fraction p_{LC}	0,8
Humidity Class	CH
Safe overload	200 % of E_{max}
Recommended excitation	12 V DC
Excitation maximum	12 V DC
Transducer material	Stainless steel
Atmospheric protection	Hermetically welded
Electromagnetic environment class	E2
Number of counts for E_{max}	$\geq Y * 5 / p_{LC}$
Software identification	Version number: XRCv2xxxxxx ⁽²⁾

Remarks:

1. The characteristics for n_{max} , Y and Z can be reduced separately.
2. xxxxxx is a number between 000000 and 999999 representing updates of the non legally relevant part of the software.

Each produced load cell is provided with an accompanying document with information about its characteristics.

The above identified Type (represented by the sample(s) identified in the OIML Test Report) have been found to comply with the additional national requirements established by the United States of America (NIST Handbook 44 and NCWM Publication 14), included in the Utilizer Declaration:

- R 60 OIML-CS rev.2 Additional requirements from the United States Accuracy class III L;
- R 60 OIML-CS rev.2 Additional requirements from the United States Marking requirements.



OIML Member State
The Netherlands

OIML Certificate of Conformity

Number R60/2000-NL1-17.58 revision 1
Project number 2845853
Page 3 of 3

Revision History

This revision replaces the previous version(s).

Revision	Date	Change(s)
Initial	2017-09-05	Initial issue
1	2021-11-30	Revision to include reports with NTEP annex