



**FEDERAL AGENCY FOR TECHNICAL REGULATION
AND METROLOGY OF RUSSIA**

OIML Member State
The Russian Federation

OIML Certificate №
R 60/2000-RU1-05.03

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: VNIIMS
Address: 46, Ozernaya str, G-361, Moscow, 119361,
the Russian Federation
Person responsible: Dr. S.A. Kononogov

Applicant

Name: PULS ELECTRONIK SISTEMLERI
MAK. SAN. IÇ VE DIŞ TIC. LTD. ŞTI.
Address: Adnan Kahveci Cad. Kos Koop.
Sanayi Sitesi 2, Blok No 9
Omerli – Hadimköy - Catalca, Istanbul, Turkey

Manufacturer of the certified pattern is the Applicant

Identification of the certified pattern: Type RT-3C

Further characteristics are set out on page 2

This certificate attests the conformity of the above-mentioned pattern (represented by the sample identified in the associated test reports) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R 60
edition 2000
for accuracy class C
 $P_i = 0,7$

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

This certificate does not bestow any form of legal international approval.

The conformity was established by tests described in associated test report number 401-05/03 which includes 26 pages.

The issuing authority

VNIIMS



Dr. S.A. Kononogov

Director

Date: 25.01.2005

The CIML member

Dr. L. K. Issaev

Professor

Date: 25.01.2005

Characteristics:

Accuracy class	C
Maximum capacity (E_{\max})	100 kg - 300 kg – 600 kg
Maximum number of load cell verification intervals (n_{\max})	3000
Ratio of minimum LC verification interval $Y = E_{\max}/V_{\min}$	10000
Temperature range	- 10 ⁰ C/+40 ⁰ C

Important note: Apart from the mention of the certificates reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.