



OIML Member State
The Netherlands

OIML Certificate

Number R60/2017-A-NL1-19.22 Revision 1
Project number 2526581
Page 1 of 3

Issuing authority	NMi Certin B.V. Person responsible: M.Ph.D. Schmidt
Applicant and Manufacturer	Flintec UK Ltd W4/5 Capital Point, Capital Business Park Wentloog Avenue Cardiff, CF3 2PW United Kingdom
Identification of the certified type	A bending beam load cell , with strain gauges. Type : SB6
Characteristics	See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 60 - Edition 2017 (E) for accuracy class C

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority	NMi Certin B.V., OIML Issuing Authority NL1 20 September 2021
-------------------	---

Certification Board

NMi Certin B.V.
Thijssseweg 11
2629 JA Delft
The Netherlands
T +31 88 6362332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.



The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- No. SN 1374 dated 30 January 2017 that includes 22 pages.

Characteristics of the load cell:

Characterization of load cell capabilities	Analog-passive load cell
Maximum capacity (E_{max})	50 kg up to and including 25 kg
Minimum dead load	0 kg
Accuracy Class	C
Rated Output	2,0 mV/V \pm 0,02 mV/V
Maximum number of load cell intervals (n) ⁽¹⁾	4000
Ratio of minimum LC Verification interval ⁽¹⁾ $Y = E_{max} / V_{min}$	22727
Ratio of minimum dead load output return ⁽¹⁾ $Z = E_{max} / (2 * DR)$	6270
Input impedance	1100 $\Omega \pm$ 50 Ω
Temperature range	-10 °C / +40 °C
Fraction p_{LC}	0,7
Humidity Class	CH
Safe overload	200 % of E_{max}
Output impedance	1000 $\Omega \pm$ 2 Ω
Recommended excitation	5 V DC to 15 V DC
Excitation maximum	15 V DC
Transducer material	Stainless steel
Atmospheric protection	Hermetic sealing (IP68 / IP69K)

Remarks:

1. The characteristics for n_{max} , Y and Z can be reduced separately.

Each load cell produced is provided with an accompanying document with information about its characteristics.



OIML Member State
The Netherlands

OIML Certificate

Number R60/2017-A-NL1-19.22 Revision 1
Project number 2526581
Page 3 of 3

Revision History

This revision replaces the previous version.

Revision	Date	Change(s)
Initial	2 September 2019	-
1	20 September 2021	Change of the identification of the certified type (deleted 'equipped with electronics')