

OIML Member State
The Netherlands

Number R60/2017-A-NL1-20.14
Project number 2480712
Page 1 of 2

Issuing authority

NMi Certin B.V.
Person responsible: M. Boudewijns

Applicant and
Manufacturer

MOSAŞ AKILLI ULAŞIM TEKNOLOJİLERİ A.Ş.
Büyükkayacak OSB Mh. 422. Sk. No:13/1
Selçuklu, Konya
Turkey

Identification of the
certified type

A **compression load cell**, with strain gauges.
Registered trade name : TARALSA
Type : TRS-M1, TRS-L1, TRS-L2

Characteristics

See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 60 - Edition 2017 (E) for accuracy class C

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

NMi Certin B.V., OIML Issuing Authority NL1
28 April 2020

Certification Board

NMi Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 6362332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.



The conformity was established by the results of tests and examinations provided in the associated OIML Test Reports:

- No. R60\2000-NL1-05.17A dated 12 October 2005 that includes 40 pages;
- No. R60\2000-NL1-05.17B dated 11 October 2005 that includes 37 pages.

Characteristics of the load cell:

Characterization of load cell capabilities	Analog-passive load cell
Maximum capacity (E_{max})	10000 kg up to and including 50000 kg
Minimum dead load	0 kg
Accuracy Class	C
Rated Output	2 mV/V \pm 0,002 mV/V
Maximum number of load cell intervals (n) ⁽¹⁾	3000
Ratio of minimum LC Verification interval ⁽¹⁾ $Y = E_{max} / V_{min}$	10000
Ratio of minimum dead load output return ⁽¹⁾ $Z = E_{max} / (2 * DR)$	3000
Input impedance	400 Ω \pm 20 Ω for TARALSA TRS-L2 730 Ω \pm 30 Ω for TARALSA TRS-L1 1100 Ω \pm 10 Ω for TARALSA TRS-M1
Temperature range	-10 °C / + 40 °C
Fraction p_{LC}	0.7
Humidity Class	CH
Safe overload	150 % of E_{max}
Output impedance	352 Ω \pm 2 Ω for TARALSA TRS-L2 705 Ω \pm 5 Ω for TARALSA TRS-L1 1005 Ω \pm 5 Ω for TARALSA TRS-M1
Recommended excitation	10 / 12 V DC
Excitation maximum	15 V DC
Transducer material	Alloy steel
Atmospheric protection	Welded steel cover

Remark:

1. The characteristics for n_{max} and Y can be reduced separately. Z is proportional or equal to n_{max} .