

OIML Member State
The Netherlands

Number R60/2017-A-NL1-20.22
Project number 2505832
Page 1 of 2

Issuing authority

NMi Certin B.V.
Person responsible: M. Boudewijns

Applicant and
Manufacturer

Coti Global Sensors, Inc.
5699 Highway 53
Harvest, AL35749
United States of America

Identification of the
certified type

A **bending beam** or **shear beam load cell**, with strain gauges.

Registered trade name : Coti Global Sensors, Inc.

Type : CG-23, CG-23-LP, CG-23-SS, CG-23-SSW, CG-743, CG-743-SSW, CG-745, CG-MK15, CG-SB3, CG-SB250, CG-SSB

Characteristics

See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 60 - Edition 2017 (E) for accuracy class C

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

NMi Certin B.V., OIML Issuing Authority NL1
26 June 2020

Certification Board

NMi Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 6362332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.



The conformity was established by the results of tests and examinations provided in the associated OIML Type Evaluation Reports:

- No. NMI-2354831-01 dated 10 September 2019 that includes 51 pages;
- No. NMI-2354831-02 dated 10 September 2019 that includes 46 pages;
- No. NMI-2354831-03 dated 10 September 2019 that includes 48 pages.

Characteristics of the load cell:

Characterization of load cell capabilities	Analog-passive load cell		
Load cell construction	Bending beam	Shear beam	
Maximum capacity (E_{max})	45,3 kg up to 453 kg	453 kg up to 2268 kg	2268 kg up to and including 11340 kg
Minimum dead load	0 kg		
Accuracy Class	C		
Rated Output	2,0 mV/V \pm 0,25% or 3,0 mV/V \pm 0,25%		
Maximum number of load cell intervals (n) ⁽¹⁾	3000	4000	
Ratio of minimum LC Verification interval ⁽¹⁾ $Y = E_{max} / V_{min}$	20000	30000	50000
Ratio of minimum dead load output return ⁽¹⁾ $Z = E_{max} / (2 * DR)$	5000	7000	8000
Input impedance	385 Ω \pm 10 Ω		
Temperature range	-10 $^{\circ}$ C / + 40 $^{\circ}$ C		
Fraction p_{LC}	0,7		
Humidity Class	CH		
Safe overload	150 % of E_{max}		
Output impedance	350 Ω \pm 3 Ω		
Recommended excitation	10 V AC / DC		
Excitation maximum	15 V AC / DC		
Transducer material	Alloy steel or Stainless steel		
Atmospheric protection	Silicon rubber and/or welded		

Remarks:

1. The characteristics for n_{max} , Y and Z can be reduced separately.

Each load cell produced is provided with an accompanying document with information about its characteristics.

The above identified Type (represented by the sample(s) identified in the OIML Test Report) have been found to comply with the additional national requirements established by the United States of America (NIST Handbook 44 and NCWM Publication 14), included in the Utilizer Declaration:

- R 60 OIML-CS rev.2 Additional requirements from the United States Accuracy class III L;
- R 60 OIML-CS rev.2 Additional requirements from the United States Marking requirements.