

Physikalisch-Technische Bundesanstalt

Braunschweig und Berlin

Member State of OIML
Germany



OIML Certificate N°
R61/1996-DE1-04.05
Revision 1

OIML CERTIFICATE OF CONFORMITY

Issuing Authority

Name: Physikalisch-Technische Bundesanstalt
Address: Bundesallee 100, 38116 Braunschweig
Person responsible: Dr. Panagiotis Zervos

Applicant

Name: Premier Tech
Address: 1 avenue Premier, G5R 6C1 Rivière-du-Loup
Canada

Manufacturer of the certified type is the applicant.

Identification of the certified type Automatic Gravimetric Filling Instrument
Type: SpeedAC NXT

Further characteristics see page 2

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R61-1, edition 1996
for accuracy class Ref (0.2)

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

Physikalisch-Technische Bundesanstalt

OIML Certificate N°
R61/1996-DE1-04.05
Revision 1

The conformity was established by the results of tests and examinations provided in the Report No. 1.12-40212666 and Test Report No. 1.12-4012666/1.

This revision has been issued because the certificate devolves to the new owner Premier Tech Canada.

The Issuing Authority

Dr. P. Zervos
Regierungsdirektor

25.06.2007

The OIML Member

Dr. R. Schwartz
Direktor und Professor

25.06.2007

Identification of the type (continued):

functioning as:	fill weigher or subtractive weigher
maximum capacities:	≤ 3000 kg
scale intervals	1 g ... 500 g
number of scale intervals	≤ 10000
Minimum fills:	depending on scale interval d, see annexed report

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report(s) is not permitted, although either may be reproduced in full.