

Member State  
P.R.China

OIML Certificate No.  
R76/1992- CN1-05.01

# OIML CERTIFICATE OF CONFORMITY

## Issuing authority

Name: Office of OIML Affairs, General Administration of Quality Supervision, Inspection and Quarantine  
Address: No.9 Madiandonglu, Haidian District, Beijing, 100088 China

## Applicant

Name: W & P Scales Mfg. (Kunshan) Co., Ltd.  
Address: No.99. Shunchang Rd. Kunshan, Jiangsu Province, 215337 P.R. China

## Manufacturer of the certified type (if the manufacturer is not the applicant)

Name:  
Address:

## Identification of the certified type

Price computing instrument


Type	QTP-30kg
Accuracy class	III
Max	30kg
Min	100g
e	5g (< 15kg = /10g( 15kg)
n	3000g
Temperature range	0-40
Humidity range	0-80%

The numbers of the associated test reports and the included pages:  
2004-58(O) 54 pages

This certificate attests the conformity of the above-mentioned type(represented by the samples identified in the associated test report) with the requirements of the following Recommendation(s) of the International Organization of Legal Metrology (OIML).

R: 76

Edition: 1992

For accuracy class (if applicable): 

This certificate relates only to the metrological and technical characteristics of the type of the instrument concerned ,as covered by the relevant OIML International Recommendation(s).

This certificate does not bestow any form of legal international approval. The conformity was established by tests described in the associated test reports.

The issuing authority  
  
Office of OIML Affairs

October 17, 2005

The CIML member

王秦平

October 17, 2005

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.